

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]	Comparability						AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
		RoM	SD	CV [%]	N _{tot}	N _{out}									
(130) WBC															
Samples and groups	f.10(exp9)/L]				21								21	21	100%
Sample A		6,29	0,30	4,8	21	CVP	6,37	0,034	15%	5,41	7,33	21	21	100%	
All results		6,29	0,30	4,8	21	0						21	21	100%	
Sample B		4,37	0,16	3,6	21	CVP	4,47	0,029	15%	3,79	5,15	21	21	100%	
All results		4,37	0,16	3,6	21	0						21	21	100%	
(160) Neutrophiles															
Samples and groups	f.10(exp9)/L]				21								21	20	95%
Sample A		3,34	0,16	4,8	21	CVP	3,31	0,023	25%	2,48	4,14	21	20	95%	
All results		3,34	0,16	4,8	21	0						21	20	95%	
Sample B		3,16	0,12	3,8	21	CVP	3,18	0,027	25%	2,38	3,98	21	21	100%	
All results		3,16	0,12	3,8	21	0						21	21	100%	
(161) Lymphocytes															
Samples and groups	f.10(exp9)/L]				21								21	19	90%
Sample A		2,28	0,17	7,2	21	CVP	2,34	0,020	25%	1,75	2,93	21	21	100%	
All results		2,28	0,17	7,2	21	0						21	21	100%	
Sample B		0,869	0,076	8,8	21	CVP	0,871	0,014	25%	0,653	1,09	21	19	90%	
All results		0,869	0,076	8,8	21	8						21	19	90%	
(162) Monocytes															
Samples and groups	f.10(exp9)/L]				21								21	21	100%
Sample A		0,545	0,086	16	21	CVP	0,571	0,013	50%	0,285	0,857	21	21	100%	
All results		0,545	0,086	16	21	0						21	21	100%	
Sample B		0,300	0,063	21	21	CVP	0,312	0,081	50%	0,156	0,468	21	21	100%	
All results		0,300	0,063	21	21	0						21	21	100%	
(163) Eosinophiles															
Samples and groups	f.10(exp9)/L]				21								21	20	95%
Sample A		0,103	0,017	16	21	CVP	0,1	0,027	asym.	0	0,192	21	20	95%	
All results		0,103	0,017	16	21	0						21	20	95%	
Sample B		0,048	0,013	28	21	CVP	0,049	0,024	asym.	0	0,135	21	21	100%	
All results		0,048	0,013	28	21	0						21	21	100%	
(164) Basophiles															
Samples and groups	f.10(exp9)/L]				21								21	21	100%
Sample A		0,020	0,018	93	21	CVP	0,026	0,027	asym.	0	0,192	21	21	100%	
All results		0,020	0,018	93	21	0						21	21	100%	
Sample B		0,030	0,021	70	21	CVP	0,04	0,027	asym.	0	0,135	21	21	100%	
All results		0,030	0,021	70	21	0						21	21	100%	