

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	3	Negative, normal, <= 2,8 mmol/L	0,96 %
		8	2,81 - 11 mmol/L	2,6 %
		> 135	11,1 - 28 mmol/L	43 %
		>>> 165	> 28 mmol/L	53 %
----- Total per sample: n = 311 ----- Success: 96 % -----				
B	All results	>>> 308	Negative, normal, <= 2,8 mmol/L	99 %
		2	11,1 - 28 mmol/L	0,64 %
		1	> 28 mmol/L	0,32 %
----- Total per sample: n = 311 ----- Success: 99 % -----				
Total success: 96 %				

**(263) Ketone bodies**

A	All results	3	Negative, <= 0,5 mmol/L	0,96 %
		> 98	1+, 2+	32 %
		>>> 210	3+, 4+	68 %
----- Total per sample: n = 311 ----- Success: 99 % -----				
B	All results	>>> 308	Negative, <= 0,5 mmol/L	99 %
		1	1+, 2+	0,32 %
		2	3+, 4+	0,64 %
----- Total per sample: n = 311 ----- Success: 99 % -----				
Total success: 99 %				

**(264) Total protein**

A	All results	3	Negative, < 0,3 g/L	0,97 %
		4	0,3 - 0,5 g/L	1,3 %
		> 60	0,51 - 2,5 g/L	19 %
		>>> 243	> 2,5 g/L	78 %
----- Total per sample: n = 310 ----- Success: 98 % -----				
B	All results	>>> 306	Negative, < 0,3 g/L	99 %
		1	0,3 - 0,5 g/L	0,32 %
		3	> 2,5 g/L	0,97 %
----- Total per sample: n = 310 ----- Success: 99 % -----				
Total success: 97 %				

**(266) Urobilinogen**

A	All results	5	Negative, < 17 µmol/L	1,6 %
		> 47	1+, 2+	15 %
		>>> 259	3+, 4+	83 %
----- Total per sample: n = 311 ----- Success: 98 % -----				
B	All results	>>> 306	Negative, < 17 µmol/L	98 %
		2	1+, 2+	0,64 %
		3	3+, 4+	0,96 %
----- Total per sample: n = 311 ----- Success: 98 % -----				
Total success: 98 %				

**(268) Bilirubin**

A	All results	1	Not specified	0,33 %
		7	Negative	2,3 %
		> 25	1+, 2+	8,2 %
		>>> 272	3+, 4+	89 %
----- Total per sample: n = 305 ----- Success: 97 % -----				
B	All results	1	Not specified	0,33 %
		>>> 300	Negative	98 %
		1	1+, 2+	0,33 %
		3	3+, 4+	0,98 %
----- Total per sample: n = 305 ----- Success: 98 % -----				
Total success: 97 %				

**(270) WBC**

A	All results	4	Negative	1,5 %
		3	10 - 50 leu/µL	1,1 %
		> 49	51 - 250 leu/µL	19 %
		>>> 206	> 250 leu/µL	79 %
----- Total per sample: n = 262 ----- Success: 97 % -----				
B	All results	>>> 259	Negative	99 %
		3	> 250 leu/µL	1,1 %
----- Total per sample: n = 262 ----- Success: 99 % -----				
Total success: 97 %				

**(272) Blood**

## Test

Sample	Group	N	Result	N <sub>rel</sub>
A	All results	4	Negative	1,3 %
		> 23	13 - 50 ery/ $\mu$ L	7,4 %
		>>> 282	> 50 ery/ $\mu$ L	91 %
Total per sample: n = 309		Success: 99 %		
B	All results	1	Not specified	0,32 %
		>>> 290	Negative	94 %
		11	5 - 12 ery/ $\mu$ L	3,6 %
		3	13 - 50 ery/ $\mu$ L	0,97 %
		4	> 50 ery/ $\mu$ L	1,3 %
Total per sample: n = 309		Success: 94 %		
Total success: 94 %				

## (273) pH

A	All results	3	5,0	0,96 %
		> 16	6,5	5,1 %
		> 55	7,0	18 %
		>>> 75	7,5	24 %
		>>> 147	8,0	47 %
		> 14	8,5	4,5 %
		> 1	9,0	0,32 %
Total per sample: n = 311		Success: 99 %		
B	All results	>>> 266	5,0	86 %
		> 32	5,5	10 %
		> 9	6,0	2,9 %
		1	6,5	0,32 %
		2	7,5	0,64 %
		1	8,0	0,32 %
Total per sample: n = 311		Success: 99 %		
Total success: 99 %				

## (274) Nitrite

A	All results	5	Negative	1,9 %
		>>> 264	Positive	98 %
Total per sample: n = 269		Success: 98 %		
B	All results	>>> 266	Negative	99 %
		3	Positive	1,1 %
Total per sample: n = 269		Success: 99 %		
Total success: 98 %				

## (275) hCG

A	All results	1	Negative	5,9 %
		>>> 16	Positive	94 %
Total per sample: n = 17		Success: 94 %		
B	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 94 %				