

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(261) Glucose				
A	All results	1	Negative, normal, <= 2,8 mmol/L	4,0 %
		>	8 11,1 - 28 mmol/L	32 %
		>>>	16 > 28 mmol/L	64 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
B	All results	>>>	24 Negative, normal, <= 2,8 mmol/L	96 %
			1 > 28 mmol/L	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 96 %				

(263) Ketone bodies

A	All results	1	Negative, <= 0,5 mmol/L	4,0 %
		>	7 1+, 2+	28 %
		>>>	17 3+, 4+	68 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
B	All results	>>>	24 Negative, <= 0,5 mmol/L	96 %
			1 3+, 4+	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 96 %				

(264) Total protein

A	All results	1	Negative, < 0,3 g/L	4,0 %
		>	7 0,51 - 2,5 g/L	28 %
		>>>	17 > 2,5 g/L	68 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
B	All results	>>>	24 Negative, < 0,3 g/L	96 %
			1 > 2,5 g/L	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 96 %				

(266) Urobilinogen

A	All results	1	Negative, < 17 µmol/L	4,0 %
		>	3 1+, 2+	12 %
		>>>	21 3+, 4+	84 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
B	All results	>>>	24 Negative, < 17 µmol/L	96 %
			1 3+, 4+	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 96 %				

(268) Bilirubin

A	All results	2	Negative	8,0 %
		>	3 1+, 2+	12 %
		>>>	20 3+, 4+	80 %
----- Total per sample: n = 25 ----- Success: 92 % -----				
B	All results	>>>	24 Negative	96 %
			1 3+, 4+	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 92 %				

(270) WBC

A	All results	1	Negative	4,3 %
		>	4 51 - 250 leu/µL	17 %
		>>>	18 > 250 leu/µL	78 %
----- Total per sample: n = 23 ----- Success: 96 % -----				
B	All results	>>>	22 Negative	96 %
			1 > 250 leu/µL	4,3 %
----- Total per sample: n = 23 ----- Success: 96 % -----				
Total success: 96 %				

(272) Blood

A	All results	1	Negative	4,0 %
		>>>	24 > 50 ery/µL	96 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
B	All results	>>>	24 Negative	96 %
			1 > 50 ery/µL	4,0 %
----- Total per sample: n = 25 ----- Success: 96 % -----				
Total success: 96 %				

(273) pH

Test

Sample	Group	N	Result	N _{rel}
A	All results	1	5,0	4,0 %
		>	3 6,5	12 %
		>	7 7,0	28 %
		>>>	2 7,5	8,0 %
		>>>	11 8,0	44 %
		>	1 8,5	4,0 %
Total per sample: n = 25		Success: 96 %		
B	All results	>>>	20 5,0	80 %
		>	3 5,5	12 %
		>	1 6,0	4,0 %
			1 7,5	4,0 %
Total per sample: n = 25		Success: 96 %		
Total success: 96 %				

(274) Nitrite

A	All results	2	Negative	8,3 %
		>>>	22 Positive	92 %
Total per sample: n = 24		Success: 92 %		
B	All results	>>>	23 Negative	96 %
			1 Positive	4,2 %
Total per sample: n = 24		Success: 96 %		
Total success: 92 %				

(275) hCG

A	All results	>>>	4 Positive	100 %
Total per sample: n = 4		Success: 100 %		
B	All results	>>>	4 Negative	100 %
Total per sample: n = 4		Success: 100 %		
Total success: 100 %				