Summary statistics - qualitative results

(Groups: measurement principle)

EQA round: FC1/19 - Calprotectin in Stool

Dead line: 08.03.2019

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group		N	Result	N_{rel}
(223) Calprotectin					
A	(2) EIA (ELISA)	>>>	31	Positive	51 %
	(3) FIA (FEIA, TRACE)	>>>	1	Positive	1,6 %
	(4) LIA (CLIA, CMIA, ECLIA)		1	Inconclusive result (borderline)	1,6 %
		>>>	6	Positive	9,8 %
	(6) IMT, IMN	>>>	5	Positive	8,2 %
	(7) HPLC, GC		1	Negative	1,6 %
		>>>	15	Positive	25 %
	(99) Another measurement principle	>>>	1	Positive	1,6 %
	Total per sample: n = 61	Success: 97 %			
B	(2) EIA (ELISA)		1	Negative	1,6 %
		>>>	30	Positive	49 %
	(3) FIA (FEIA, TRACE)	>>>	1	Positive	1,6 %
	(4) LIA (CLIA, CMIA, ECLIA)	>>>	7	Positive	11 %
	(6) IMT, IMN	>>>	5	Positive	8,2 %
	(7) HPLC, GC		1	Negative	1,6 %
		>>>	15	Positive	25 %
	(99) Another measurement principle	>>>	1	Positive	1,6 %
	Total per sample: n = 61	Success: 97 %			

Total success: 95 %

st_kl

End of report

Printed: 20.03.2019