

Summary statistics - qualitative results

(Groups: measurement principle)

EQA round: HSI/19 - Haemocoagulation Special

Dead line: 24.05.2019

Key: >>> ... marks correct (expected) result

N ... number of the results

> ... marks conditionally correct (acceptable) result

Nrel ... relative number of the results

± ... marks the result not evaluated

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Set of samples 3**(488) APC - resistance**

A3	(1) Coagulometer	>>> 60	Positive	97 %
	(2) Chromogenous substrat	>>> 2	Positive	3,2 %

 Total per sample: n = 62 Success: 100 %

Total success: 100 %

st_kl

End of report

Printed: 05.06.2019