

EQA round: HSI/19 - Haemocoagulation Special

Dead line: 24.05.2019

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Set of samples 3

(488) APC - resistance

A3	(38) IL	>>>	13	Positive	21 %
	(51) Technoclone	>>>	2	Positive	3,2 %
	(94) Stago	>>>	3	Positive	4,8 %
	(102) Chromogenix	>>>	2	Positive	3,2 %
	(149) Siemens (Dade)	>>>	14	Positive	23 %
	(159) Grifols	>>>	1	Positive	1,6 %
	(192) Pentapharm	>>>	24	Positive	39 %
	(193) HYPHEN BioMed	>>>	2	Positive	3,2 %
	(999) another manufacturer	>>>	1	Positive	1,6 %

----- Total per sample: n = 62 ----- Success: 100 % -----

Total success: 100 %

st_kl

End of report

Printed: 05.06.2019