

Filtr: minimální četnost skupin n = 5

Cyklus EHK: HSI/19 - Hemokoagulace speciální

Stop termín: 24.05.2019

Zkouška	[jednotka]	RoM	SD	CV [%]	N _{tot}	N _{out}	Srovnatelnost						N _{eva}	N _{suc}	S _{rel}
							AV	U _{AV}	D _{max}	LL	UL				
(486) Protein S					73								73	73	100%
Vzorky a skupiny	[%]														
Vzorek A3		79,5	13	16	73		CVP	79,5	3,7	30%	55,6	104	73	73	100%
(149) Siemens (Dade)		79,3	16	20	35	0							35		
(193) HYPHEN BioMed		80,7	9,1	11	12	0							12		
Ostatní					5	0							5		
								2x 175, 1x 179, 1x 184, 1x 999							
(487) ProC Global (norm. poměr)					30								30	29	97%
Vzorky a skupiny	[-]														
Vzorek A3		0,458	0,047	10	30		CVP	0,458	0,021	25%	0,343	0,573	30	29	97%
(149) Siemens (Dade)		0,460	0,046	10	29	0							29		
Ostatní					1	0							1		
								1x 179							
Sada vzorků 4															
(491) Faktor II					32								32	31	97%
Vzorky a skupiny	[%]														
Vzorek A4		96,6	10	11	32		CVP	96,6	4,5	25%	72,4	121	32	31	97%
(38) IL		92,6	9,7	10	5	0							5		
(149) Siemens (Dade)		98,1	9,4	9,6	20	0							20		
Ostatní					7	0							7		
								1x 63, 3x 94, 1x 100, 2x 193							
(492) Faktor V					34								34	34	100%
Vzorky a skupiny	[%]														
Vzorek A4		95,1	5,9	6,2	34		CVP	95,1	2,5	25%	71,3	119	34	34	100%
(38) IL		92,6	3,0	3,2	5	0							5		
(94) Stago		94,5	6,7	7,1	6	0							6		
(149) Siemens (Dade)		96,4	5,8	6,0	20	0							20		
Ostatní					3	0							3		
								1x 63, 1x 100, 1x 999							
(493) Faktor VII					36								36	36	100%
Vzorky a skupiny	[%]														
Vzorek A4		115	12	10	36		CVP	115	4,8	25%	86,2	144	36	36	100%
(38) IL		102	7,2	7,0	5	0							5		
(149) Siemens (Dade)		117	12	10	20	0							20		
Ostatní					11	0							11		
								1x 63, 2x 94, 1x 100, 2x 159, 4x 193, 1x 999							
(494) Faktor X					30								30	29	97%
Vzorky a skupiny	[%]														
Vzorek A4		89,8	8,0	8,9	30		CVP	89,8	3,6	25%	67,3	113	30	29	97%
(38) IL		84,6	5,9	7,0	5	0							5		
(149) Siemens (Dade)		91,3	6,5	7,1	18	0							18		
Ostatní					7	0							7		
								1x 63, 3x 94, 3x 193							

st_kn_p

Konec sestavy

Vytiskeno: 05.06.2019