

Summary statistics - qualitative results

(Groups: manufacturer of instrument)

EQA round: IFT1/19 - Immunophenotypisation

Dead line: 15.04.2019

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(259) HLA-B27				
A	(12) Beckman Coulter	>>> 44	Negative	65 %
		1	Inconclusive result (borderline)	1,5 %
		1	Positive	1,5 %
	(88) Becton Dickinson	>>> 18	Negative	26 %
		1	Inconclusive result (borderline)	1,5 %
	(177) MINDRAY	>>> 3	Negative	4,4 %
----- Total per sample: n = 68		----- Success: 96 %	-----	
B	(12) Beckman Coulter	2	Negative	2,9 %
		> 3	Inconclusive result (borderline)	4,4 %
		>>> 41	Positive	60 %
	(88) Becton Dickinson	1	Negative	1,5 %
		> 1	Inconclusive result (borderline)	1,5 %
		>>> 17	Positive	25 %
	(177) MINDRAY	>>> 3	Positive	4,4 %
----- Total per sample: n = 68		----- Success: 96 %	-----	
Total success: 94 %				

st_kl

End of report

Printed: 21.04.2019