

## Summary statistics - qualitative results

(Groups: manufacturer of kit)

## EQA round: IFT1/19 - Immunophenotypisation

Dead line: 15.04.2019

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

## Test

Sample	Group	N	Result	N <sub>rel</sub>	
<b>(259) HLA-B27</b>					
A	(12) Beckman Coulter	>>>	40 Negative	59 %	
			2 Inconclusive result (borderline)	2,9 %	
	(36) Immunotech	>>>	1 Negative	1,5 %	
			1 Positive	1,5 %	
	(88) Becton Dickinson	>>>	20 Negative	29 %	
	(177) MINDRAY	>>>	3 Negative	4,4 %	
	(190) Exbio Praha	>>>	1 Negative	1,5 %	
	Total per sample: n = 68		Success: 96 %		
	B	(12) Beckman Coulter		3 Negative	4,4 %
			>	3 Inconclusive result (borderline)	4,4 %
(36) Immunotech		>>>	36 Positive	53 %	
			2 Positive	2,9 %	
(88) Becton Dickinson		>	1 Inconclusive result (borderline)	1,5 %	
(177) MINDRAY		>>>	19 Positive	28 %	
			3 Positive	4,4 %	
(190) Exbio Praha		>>>	1 Positive	1,5 %	
Total per sample: n = 68		Success: 96 %			
Total success: 94 %					

st\_kl

End of report

Printed: 21.04.2019