

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(253) CD 3+ T lymphocytes					31							31	29	94%
Samples and groups	[%]													
Sample A		57,9	1,2	2,1	31	CVP	58	0,36	7%	53,9	62,1	31	30	97%
All results		57,9	1,2	2,1	31	0						31		
Sample B		77,6	2,1	2,7	31	CVP	77,7	0,54	7%	72,2	83,2	31	30	97%
All results		77,6	2,1	2,7	31	0						31		
(254) CD 4+ T lymphocytes					31							31	29	94%
Samples and groups	[%]													
Sample A		49,5	1,5	3,1	31	CVP	49,8	0,38	10%	44,8	54,8	31	29	94%
All results		49,5	1,5	3,1	31	0						31		
Sample B		43,6	1,6	3,7	31	CVP	44,2	0,42	10%	39,7	48,7	31	30	97%
All results		43,6	1,6	3,7	31	0						31		
(255) CD 8+ T lymphocytes					31							31	29	94%
Samples and groups	[%]													
Sample A		7,30	0,53	7,3	31	CVP	7,3	0,13	17%	6,05	8,55	31	29	94%
All results		7,30	0,53	7,3	31	0						31		
Sample B		29,1	1,8	6,1	31	CVP	29,4	0,41	17%	24,4	34,4	31	31	100%
All results		29,1	1,8	6,1	31	0						31		
(256) B lymphocytes					31							31	29	94%
Samples and groups	[%]													
Sample A		19,4	0,86	4,4	31	CVP	19,4	0,26	30%	13,5	25,3	31	31	100%
All results		19,4	0,86	4,4	31	0						31		
Sample B		11,5	0,83	7,2	31	CVP	11,3	0,21	30%	7,91	14,7	31	29	94%
All results		11,5	0,83	7,2	31	0						31		
(257) NK cells					31							31	29	94%
Samples and groups	[%]													
Sample A		21,2	2,0	9,4	31	CVP	21,1	0,37	26%	15,6	26,6	31	29	94%
All results		21,2	2,0	9,4	31	0						31		
Sample B		9,80	1,1	11	31	CVP	9,54	0,32	26%	7,05	12,1	31	30	97%
All results		9,80	1,1	11	31	0						31		