

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

## EQA round: KD3/19 - Glycated Haemoglobin

Dead line: 30.08.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability													
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>	
<b>(220) Haemoglobin A1c</b>					176							176	162	92%	
Samples and groups	[mmol/mol]														
<b>Sample A</b>		33,4	2,0	6,0	176	CVP	33,4	0,37	12%	29,3	37,5	176	166	94%	
(14) Bio-Rad		34,1	1,8	5,3	47	0						47			
(25) TOSOH		33,7	1,5	4,3	62	0						62			
(60) Roche		36,8	4,1	11	7	0						7			
(150) Arkray		32,0	1,9	5,8	48	0						48			
Other					12	0						12			
							2x 1, 2x 34, 1x 46, 1x 49, 1x 58, 1x 75, 1x 149, 1x 177, 1x 196, 1x 999								
<b>Sample B</b>		73,7	3,4	4,6	176	CVP	73,7	0,63	12%	64,8	82,6	176	169	96%	
(14) Bio-Rad		76,1	2,7	3,5	47	0						47			
(25) TOSOH		74,1	2,0	2,7	62	0						62			
(60) Roche		77,2	2,5	3,3	7	0						7			
(150) Arkray		70,5	2,9	4,2	48	1						48			
Other					12	0						12			
							2x 1, 2x 34, 1x 46, 1x 49, 1x 58, 1x 75, 1x 149, 1x 177, 1x 196, 1x 999								

st\_kn\_p

End of report

Printed: 07.09.2019