

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

## EQA round: LMWH1/19 - Low Molecular Weight Heparin

Dead line: 24.05.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(103) Low Molecular Weight Heparin</b>					114							114	108	95%
Samples and groups	[IU/mL]													
<b>Sample A</b>		0,940	0,097	10	114	CVP	0,94	0,022	25%	0,705	1,18	114	112	98%
(38) IL		0,792	0,034	4,3	18							18		
(94) Stago		0,799	0,064	8,0	10							10		
(149) Siemens (Dade)		0,972	0,042	4,4	32							32		
(193) HYPHEN BioMed		0,989	0,056	5,7	49							49		
Other					5							5		
							3x 102, 2x 999							
<b>Sample B</b>		1,61	0,22	13	114	CVP	1,61	0,049	25%	1,2	2,02	114	109	96%
(38) IL		1,35	0,05	3,7	18							18		
(94) Stago		1,47	0,081	5,5	10							10		
(149) Siemens (Dade)		1,61	0,16	10	32							32		
(193) HYPHEN BioMed		1,73	0,14	8,0	49							49		
Other					5							5		
							3x 102, 2x 999							

st\_kn\_p

End of report

Printed: 05.06.2019