

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of groups n = 5

## EQA round: LMWH1/19 - Low Molecular Weight Heparin

Dead line: 24.05.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	$U_{AV}$ = expanded uncertainty of the assigned value ( $k = 2$ )	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	$U_{AV}$	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(103) Low Molecular Weight Heparin</b>					12							12	12	100%
Samples and groups	[IU/mL]													
<b>Sample A</b>		0,967	0,088	9,1	12	CVP	0,94	0,022	25%	0,705	1,18	12	12	100%
(193) HYPHEN BioMed		0,990	0,074	7,5	8							8		
Other					4							4		
						2x 38, 2x 149								
<b>Sample B</b>		1,59	0,18	11	12	CVP	1,61	0,049	25%	1,2	2,02	12	12	100%
(193) HYPHEN BioMed		1,67	0,13	7,6	8							8		
Other					4							4		
						2x 38, 2x 149								

st\_kn\_p

End of report

Printed: 05.06.2019