

Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: BIL2/20 - Bilirubin Neonatal

Deadline: 21.08.2020

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(215) Bilirubin total					34							34	31	91%
Samples and groups	[µmol/L]													
Sample A		374	27	7,1	34	CVP	367	5,2	20%	293	441	34	32	94%
(2) DCA, DPD		371	26	6,9	23							23		
Other					11							11		
						4x 1, 1x 3, 1x 5, 1x 6, 4x 7								
Sample B		228	11	5,0	34	CVP	224	3,1	20%	179	269	34	33	97%
(2) DCA, DPD		228	9,8	4,3	23							23		
Other					11							11		
						4x 1, 1x 3, 1x 5, 1x 6, 4x 7								

st_kn_p

End of report

Printed: 26.08.2020