

Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: CRP3/20 - C-Reactive Protein

Deadline: 21.08.2020

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(100) C-reactive protein					53							53	53	100%
----- Samples and groups -----	[mg/L]													
Sample A		23,1	1,6	7,0	53	CVP	23,4	0,31	24%	17,7	29,1	53	53	100%
(2) Immunoturbidimetry		23,1	1,7	7,4	48							48		
Other					5							5		
						1x 0, 4x 3								
Sample B		35,4	2,4	6,8	53	CVP	35,8	0,38	24%	27,2	44,4	53	53	100%
(2) Immunoturbidimetry		35,4	2,5	7,1	48							48		
Other					5							5		
						1x 0, 4x 3								

st_kn_p

End of report

Printed: 25.08.2020