

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	>>> 28	Negative, normal, <= 2,8 mmol/L	100 %
Total per sample: n = 28		Success: 100 %		
B	All results	> 2	2,81 - 11 mmol/L	7,1 %
		>>> 13	11,1 - 28 mmol/L	46 %
		>>> 13	> 28 mmol/L	46 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				
<b>(263) Ketone bodies</b>				
A	All results	>>> 28	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 28		Success: 100 %		
B	All results	> 7	1+, 2+	25 %
		>>> 21	3+, 4+	75 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				
<b>(264) Total protein</b>				
A	All results	>>> 28	Negative, < 0,3 g/L	100 %
Total per sample: n = 28		Success: 100 %		
B	All results	1	Negative, < 0,3 g/L	3,6 %
		> 8	0,51 - 2,5 g/L	29 %
		>>> 19	> 2,5 g/L	68 %
Total per sample: n = 28		Success: 96 %		
Total success: 96 %				
<b>(266) Urobilinogen</b>				
A	All results	>>> 28	Negative, < 17 µmol/L	100 %
Total per sample: n = 28		Success: 100 %		
B	All results	> 5	1+, 2+	18 %
		>>> 23	3+, 4+	82 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				
<b>(268) Bilirubin</b>				
A	All results	>>> 28	Negative	100 %
Total per sample: n = 28		Success: 100 %		
B	All results	2	Negative	7,1 %
		> 4	1+, 2+	14 %
		>>> 22	3+, 4+	79 %
Total per sample: n = 28		Success: 93 %		
Total success: 93 %				
<b>(270) WBC</b>				
A	All results	>>> 26	Negative	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	1	Negative	3,8 %
		> 1	10 - 50 leu/µL	3,8 %
		>>> 10	51 - 250 leu/µL	38 %
		>>> 14	> 250 leu/µL	54 %
Total per sample: n = 26		Success: 96 %		
Total success: 96 %				
<b>(272) Blood</b>				
A	All results	>>> 27	Negative	96 %
		1	5 - 12 ery/µL	3,6 %
Total per sample: n = 28		Success: 96 %		
B	All results	> 2	13 - 50 ery/µL	7,1 %
		>>> 26	> 50 ery/µL	93 %
Total per sample: n = 28		Success: 100 %		
Total success: 96 %				
<b>(273) pH</b>				
A	All results	>>> 23	5,0	82 %
		> 2	5,5	7,1 %
		> 3	6,0	11 %
Total per sample: n = 28		Success: 100 %		
B	All results	> 2	6,0	7,1 %
		> 6	6,5	21 %
		>>> 13	7,0	46 %

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
B	All results	>	7 7,5	25 %
Total per sample: n = 28		Success: 100 %		
Total success: 100 %				

**(274) Nitrite**

A	All results	>>>	26 Negative	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	>>>	26 Positive	100 %
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				

**(275) hCG**

A	All results	>>>	6 Negative	100 %
Total per sample: n = 6		Success: 100 %		
B	All results	>>>	6 Positive	100 %
Total per sample: n = 6		Success: 100 %		
Total success: 100 %				