

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

## Test

Sample	Group	N	Result	N <sub>rel</sub>
--------	-------	---	--------	------------------

## Set 1

## (575) Anti-thyroglobulin

A1	(1) Fluorescent methods	>>>	2	Positive	1,1 %	
	(2) EIA methods	>>>	10	Positive	5,5 %	
	(4) RIA, IRMA, RRA	>>>	1	Positive	0,55 %	
	(5) LIA, ILMA	>>>	167	Positive	92 %	
	(99) Another measurement principle	>>>	1	Positive	0,55 %	
Total per sample: n = 181		Success:	100 %	-		
B1	(1) Fluorescent methods	>>>	2	Negative	1,1 %	
	(2) EIA methods	>>>	8	Negative	4,4 %	
				1	Positive	0,55 %
	(2) EIA methods; (179) Siemens	±	1	Inconclusive result (borderline)	0,55 %	
	(4) RIA, IRMA, RRA	>>>	1	Negative	0,55 %	
	(5) LIA, ILMA; (1) Abbott	±	9	Negative	5,0 %	
			±	13	Inconclusive result (borderline)	7,2 %
			±	35	Positive	19 %
	(5) LIA, ILMA	>>>	82	Negative	45 %	
				1	Inconclusive result (borderline)	0,55 %
				2	Positive	1,1 %
		(5) LIA, ILMA; (162) Siemens (Atellica)	±	3	Negative	1,7 %
			±	9	Positive	5,0 %
(5) LIA, ILMA; (179) Siemens	±	1	Negative	0,55 %		
		±	3	Inconclusive result (borderline)	1,7 %	
		±	9	Positive	5,0 %	
(99) Another measurement principle	>>>	1	Negative	0,55 %		
Total per sample: n = 181		Success (only evaluated results):	96 %	-		
Total success (only evaluated tests): 96 %						

## (576) Anti-microsomes (TPO)

A1	(1) Fluorescent methods	>>>	2	Positive	1,1 %	
	(2) EIA methods	>>>	11	Positive	5,9 %	
	(5) LIA, ILMA			1	Negative	0,54 %
		>>>	171	Positive	92 %	
(99) Another measurement principle	>>>	1	Positive	0,54 %		
Total per sample: n = 186		Success:	99 %	-		
B1	(1) Fluorescent methods	>>>	2	Negative	1,1 %	
	(2) EIA methods	>>>	11	Negative	5,9 %	
	(5) LIA, ILMA	>>>	172	Negative	92 %	
	(99) Another measurement principle	>>>	1	Negative	0,54 %	
Total per sample: n = 186		Success:	100 %	-		
Total success: 99 %						

## Set 2

## (590) Anti-TSH receptor

A2	(2) EIA methods	>>>	5	Negative	5,3 %
	(4) RIA, IRMA, RRA	>>>	3	Negative	3,2 %
	(5) LIA, ILMA	>>>	87	Negative	92 %
Total per sample: n = 95		Success:	100 %	-	
B2	(2) EIA methods	>>>	5	Positive	5,3 %
	(4) RIA, IRMA, RRA	>>>	3	Positive	3,2 %
	(5) LIA, ILMA	>>>	87	Positive	92 %
Total per sample: n = 95		Success:	100 %	-	
Total success: 100 %					