

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

## EQA round: AT2/21 - Autoantibodies in Thyroid Diseases

Deadline: 01.11.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>Set 2</b>														
<b>(590) Anti-TSH receptor</b>					10						10	10	100%	
Samples and groups	[U/L]													
<b>Sample A2</b>		0,750	0,09	12	10	CVP	0,789	,0060	100%	0	1,58	10	10	100%
(60) Roche		0,790	0,015	1,9	8	0						8		
Other					2	0						2		
							1x 24, 1x 29							
<b>Sample B2</b>		5,97	0,58	9,7	10	CVP	5,9	0,13	33%	3,95	7,85	10	10	100%
(60) Roche		5,74	0,73	13	8	0						8		
Other					2	0						2		
							1x 24, 1x 29							

st\_kn\_p

End of report

Printed: 04.11.2021