

## Summary statistics - quantitative results

(Groups: system)

Filter: minimal size of the groups n = 5

## EQA round: CRPP5/21 - C-Reactive Protein POCT

Deadline: 22.10.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(104) C-reactive protein</b>					261							261	224	86%
Samples and groups	[mg/L]													
<b>Sample A</b>		60,8	5,8	9,5	261	CVP	60,8	0,87	24%	46,2	75,4	261	237	91%
(130) Aidian QuikRead 101		60,6	5,0	8,2	40	0						40		
(131) Axis-Shield NycoCard		50,0	15	29	14	0						14		
(133) BodiTech i-CHROMA		60,8	18	30	15	0						15		
(134) Aidian QuikRead go		61,5	4,6	7,5	176	0						176		
(135) Abbott Alere Afinion		55,5	0,74	1,3	5	0						5		
(999) another system		58,6	9,7	17	9	0						9		
Other					2	0						2		
							1x 132, 1x 136							
<b>Sample B</b>		29,6	2,8	9,6	261	CVP	29,6	0,43	24%	22,4	36,8	261	230	88%
(130) Aidian QuikRead 101		28,2	1,8	6,3	40	0						40		
(131) Axis-Shield NycoCard		25,1	9,0	36	14	1						14		
(133) BodiTech i-CHROMA		37,4	15	39	15	0						15		
(134) Aidian QuikRead go		29,8	2,3	7,7	176	0						176		
(135) Abbott Alere Afinion		27,5	0,74	2,7	5	0						5		
(999) another system		28,9	3,2	11	9	0						9		
Other					2	0						2		
							1x 132, 1x 136							

st\_kn\_p

End of report

Printed: 30.10.2021