

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
(261) Glucose				
A	All results	>>> 24	Negative, normal, <= 2,8 mmol/L	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	>>> 4	2,81 - 11 mmol/L	17 %
		>>> 10	11,1 - 28 mmol/L	42 %
		>>> 10	> 28 mmol/L	42 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(263) Ketone bodies				
A	All results	>>> 24	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	> 4	1+, 2+	17 %
		>>> 20	3+, 4+	83 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(264) Total protein				
A	All results	>>> 24	Negative, < 0,3 g/L	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	> 6	0,51 - 2,5 g/L	25 %
		>>> 18	> 2,5 g/L	75 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(266) Urobilinogen				
A	All results	>>> 24	Negative, < 17 µmol/L	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	> 11	1+, 2+	46 %
		>>> 13	3+, 4+	54 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(268) Bilirubin				
A	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
B	All results (without individual groups)	>>> 19	3+, 4+	79 %
	(223) Iris iChem Velocity	± 2	Negative	8,3 %
		± 2	1+, 2+	8,3 %
		± 1	3+, 4+	4,2 %
Total per sample: n = 24		Success (only evaluated results): 100 %		
Total success (only evaluated tests): 100 %				
(270) WBC				
A	All results	>>> 22	Negative	100 %
Total per sample: n = 22		Success: 100 %		
B	All results	> 2	10 - 50 leu/µL	9,1 %
		>>> 6	51 - 250 leu/µL	27 %
		>>> 14	> 250 leu/µL	64 %
Total per sample: n = 22		Success: 100 %		
Total success: 100 %				
(272) Blood				
A	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	>>> 24	> 50 ery/µL	100 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(273) pH				
A	All results	>>> 18	5,0	75 %
		> 3	5,5	13 %
		> 3	6,0	13 %
Total per sample: n = 24		Success: 100 %		
B	All results	> 2	6,0	8,3 %
		> 8	6,5	33 %
		>>> 12	7,0	50 %
		> 2	7,5	8,3 %
Total per sample: n = 24		Success: 100 %		

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Total success: 100 %

(274) Nitrite

A	All results	>>> 22	Negative	100 %
Total per sample: n = 22		Success: 100 %		
B	All results	>>> 22	Positive	100 %
Total per sample: n = 22		Success: 100 %		

Total success: 100 %

(275) hCG

A	All results	>>> 5	Negative	100 %
Total per sample: n = 5		Success: 100 %		
B	All results	>>> 5	Positive	100 %
Total per sample: n = 5		Success: 100 %		

Total success: 100 %

st_kl

End of report

Printed: 10.12.2021