

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}	
(223) Calprotectin					
A	(2) EIA (ELISA)	>>>	21 Negative	31 %	
	(3) FIA (FEIA, TRACE)	>>>	4 Negative	5,9 %	
			1 Positive	1,5 %	
	(4) LIA (CLIA, CMIA, ECLIA, LOCI)	>>>	12 Negative	18 %	
	(6) IMT, IMN	>>>	13 Negative	19 %	
			1 Positive	1,5 %	
	(7) HPLC, GC	>>>	14 Negative	21 %	
	(99) Another measurement principle	>>>	2 Negative	2,9 %	
	----- Total per sample: n = 68 -----		Success: 97 %		
	B	(2) EIA (ELISA)	>>>	16 Negative	24 %
>			2 Inconclusive result (borderline)	2,9 %	
			3 Positive	4,4 %	
(3) FIA (FEIA, TRACE)		>>>	3 Negative	4,4 %	
			2 Positive	2,9 %	
(4) LIA (CLIA, CMIA, ECLIA, LOCI)		>>>	8 Negative	12 %	
		>	2 Inconclusive result (borderline)	2,9 %	
			2 Positive	2,9 %	
(6) IMT, IMN		>>>	13 Negative	19 %	
		>	1 Inconclusive result (borderline)	1,5 %	
(7) HPLC, GC		>>>	11 Negative	16 %	
		>	1 Inconclusive result (borderline)	1,5 %	
			2 Positive	2,9 %	
(99) Another measurement principle	>>>	2 Negative	2,9 %		
----- Total per sample: n = 68 -----		Success: 87 %			
Total success: 85 %					