

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of the groups n = 5

EQA round: HKG4/21 - Haemocoagulation Tests

Deadline: 19.11.2021

| | | |
|--|--|--|
| RoM = robust average | AV = assigned value | Dmax = acceptable percent difference |
| SD = standard deviation | CRV = certified reference value | LL = lower limit |
| CV = coefficient of variation | RV = reference value | UL = upper limit |
| Ntot = total number of the participants | CVE = consensus value from experts | Neva = number of evaluated participants |
| Nout = number of results excluded before calculation | CVP = consensus value from all participants | Nsuc = number of successful participants |
| | CVPG = consensus value from participants groups | Srel = success (relative) |
| | U _{AV} = expanded uncertainty of the assigned value (k = 2) | |

| Test | [unit] | Comparability | | | | | Comparability | | | | | | | | |
|---|--------|---------------|-------|--------|------------------|------------------|---------------|--|------------------|-----|-------|------------------|------------------|------------------|-----|
| | | RoM | SD | CV [%] | N _{tot} | N _{out} | AV | U _{AV} | D _{max} | LL | UL | N _{eva} | N _{suc} | S _{rel} | |
| (170) APTT - ratio | | | | | 306 | | | | | | 295 | 288 | 98% | | |
| — Samples and groups | [-] | | | | | | | | | | | | | | |
| Sample A | | 1,10 | 0,084 | 7,7 | 306 | | | | | | 295 | 294 | 100% | | |
| (38) IL; (253) IL APTT-SP | | 0,951 | 0,052 | 5,4 | 22 | 0 | CVPG | 0,952 | 0,026 | 20% | 0,761 | 1,15 | 22 | | |
| (38) IL; (256) IL SynthASil | | 1,13 | 0,049 | 4,3 | 11 | 0 | CVPG | 1,13 | 0,036 | 20% | 0,904 | 1,36 | 11 | | |
| (94) Stago; (260) Stago PTT Automate | | 1,05 | 0,043 | 4,1 | 16 | 0 | CVPG | 1,05 | 0,026 | 20% | 0,84 | 1,26 | 16 | | |
| (94) Stago; (265) Stago Cephascreen | | 1,05 | 0,049 | 4,6 | 16 | 0 | CVPG | 1,05 | 0,030 | 20% | 0,84 | 1,26 | 16 | | |
| (149) Siemens (Dade, BN, Dimension); (240) Siemens (Dade) Actin FS | | 1,14 | 0,05 | 4,4 | 92 | 0 | CVPG | 1,14 | 0,013 | 20% | 0,912 | 1,37 | 92 | | |
| (149) Siemens (Dade, BN, Dimension); (241) Siemens (Dade) Actin FSL | | 1,01 | 0,036 | 3,5 | 27 | 0 | CVPG | 1,01 | 0,017 | 20% | 0,808 | 1,22 | 27 | | |
| (149) Siemens (Dade, BN, Dimension); (242) Siemens (Dade) Pathromtin SL | | 1,15 | 0,044 | 3,9 | 79 | 0 | CVPG | 1,15 | 0,012 | 20% | 0,92 | 1,38 | 79 | | |
| (175) DIAGON; (257) Diagon Dia-PTT-Liquid | | 1,09 | 0,037 | 3,4 | 9 | 0 | CVPG | 1,09 | 0,046 | 20% | 0,872 | 1,31 | 9 | | |
| (177) Mindray; (246) Mindray APTT Reagent | | 0,995 | 0,037 | 3,7 | 9 | 0 | CVPG | 0,995 | 0,046 | 20% | 0,796 | 1,2 | 9 | | |
| (193) HYPHEN BioMed; (266) Hyphen BioMed CEPHEN | | 1,12 | 0,037 | 3,3 | 9 | 0 | CVPG | 1,12 | 0,046 | 20% | 0,896 | 1,35 | 9 | | |
| Other | | | | | 16 | 0 | | | | | | | 5 | | |
| | | | | | | | | 1x 38/999, 1x 46/999, 1x 51/253, 1x 51/270, 1x 63/242, 2x 94/259, 1x 166/248, 4x 175/269, 1x 177/999, 2x 179/240, 1x 179/242 | | | | | | | |
| Sample B | | 2,45 | 0,44 | 18 | 306 | | | | | | 295 | 289 | 98% | | |
| (38) IL; (253) IL APTT-SP | | 2,09 | 0,12 | 5,7 | 22 | 0 | CVPG | 2,09 | 0,059 | 20% | 1,67 | 2,51 | 22 | | |
| (38) IL; (256) IL SynthASil | | 2,04 | 0,12 | 5,8 | 11 | 0 | CVPG | 2,04 | 0,087 | 20% | 1,63 | 2,45 | 11 | | |
| (94) Stago; (260) Stago PTT Automate | | 2,23 | 0,088 | 3,9 | 16 | 0 | CVPG | 2,23 | 0,054 | 20% | 1,78 | 2,68 | 16 | | |
| (94) Stago; (265) Stago Cephascreen | | 2,11 | 0,13 | 6,3 | 16 | 0 | CVPG | 2,11 | 0,081 | 20% | 1,68 | 2,54 | 16 | | |
| (149) Siemens (Dade, BN, Dimension); (240) Siemens (Dade) Actin FS | | 2,37 | 0,13 | 5,6 | 92 | 0 | CVPG | 2,37 | 0,034 | 20% | 1,89 | 2,85 | 92 | | |
| (149) Siemens (Dade, BN, Dimension); (241) Siemens (Dade) Actin FSL | | 1,88 | 0,12 | 6,6 | 27 | 0 | CVPG | 1,88 | 0,058 | 20% | 1,5 | 2,26 | 27 | | |
| (149) Siemens (Dade, BN, Dimension); (242) Siemens (Dade) Pathromtin SL | | 3,00 | 0,16 | 5,2 | 79 | 1 | CVPG | 3,0 | 0,043 | 20% | 2,4 | 3,6 | 79 | | |
| (175) DIAGON; (257) Diagon Dia-PTT-Liquid | | 2,62 | 0,20 | 7,7 | 9 | 0 | CVPG | 2,62 | 0,25 | 20% | 2,09 | 3,15 | 9 | | |
| (177) Mindray; (246) Mindray APTT Reagent | | 2,32 | 0,10 | 4,5 | 9 | 0 | CVPG | 2,32 | 0,13 | 20% | 1,85 | 2,79 | 9 | | |
| (193) HYPHEN BioMed; (266) Hyphen BioMed CEPHEN | | 2,63 | 0,037 | 1,4 | 9 | 0 | CVPG | 2,63 | 0,046 | 20% | 2,1 | 3,16 | 9 | | |
| Other | | | | | 16 | 0 | | | | | | | 5 | | |
| | | | | | | | | 1x 38/999, 1x 46/999, 1x 51/253, 1x 51/270, 1x 63/242, 2x 94/259, 1x 166/248, 4x 175/269, 1x 177/999, 2x 179/240, 1x 179/242 | | | | | | | |
| (171) Fibrinogen | | | | | 237 | | | | | | 237 | 229 | 97% | | |
| — Samples and groups | [g/L] | | | | | | | | | | | | | | |
| Sample A | | 3,25 | 0,22 | 6,8 | 237 | | CVP | 3,25 | 0,035 | 25% | 2,43 | 4,07 | 237 | 233 | 98% |
| (38) IL | | 3,27 | 0,19 | 5,8 | 30 | 0 | | | | | | | 30 | | |
| (94) Stago | | 3,30 | 0,15 | 4,7 | 27 | 0 | | | | | | | 27 | | |
| (149) Siemens (Dade, BN, Dimension) | | 3,28 | 0,22 | 6,7 | 151 | 0 | | | | | | | 151 | | |
| (175) DIAGON | | 2,95 | 0,25 | 8,5 | 12 | 0 | | | | | | | 12 | | |
| (177) Mindray | | 3,01 | 0,31 | 10 | 10 | 0 | | | | | | | 10 | | |
| Other | | | | | 7 | 0 | | | | | | | 7 | | |
| | | | | | | | | 1x 46, 1x 51, 2x 179, 3x 193 | | | | | | | |
| Sample B | | 1,43 | 0,14 | 10 | 237 | | CVP | 1,43 | 0,023 | 25% | 1,07 | 1,79 | 237 | 231 | 97% |
| (38) IL | | 1,53 | 0,098 | 6,4 | 30 | 0 | | | | | | | 30 | | |
| (94) Stago | | 1,56 | 0,08 | 5,1 | 27 | 0 | | | | | | | 27 | | |
| (149) Siemens (Dade, BN, Dimension) | | 1,37 | 0,11 | 8,1 | 151 | 0 | | | | | | | 151 | | |
| (175) DIAGON | | 1,44 | 0,099 | 6,9 | 12 | 0 | | | | | | | 12 | | |
| (177) Mindray | | 1,46 | 0,12 | 8,2 | 10 | 0 | | | | | | | 10 | | |
| Other | | | | | 7 | 0 | | | | | | | 7 | | |
| | | | | | | | | 1x 46, 1x 51, 2x 179, 3x 193 | | | | | | | |
| (177) Antithrombin | | | | | 171 | | | | | | 171 | 162 | 95% | | |
| — Samples and groups | [%] | | | | | | | | | | | | | | |
| Sample A | | 106 | 5,1 | 4,9 | 171 | | CVP | 106 | 0,96 | 18% | 86,9 | 126 | 171 | 170 | 99% |
| (38) IL | | 106 | 5,2 | 4,9 | 29 | 0 | | | | | | | 29 | | |
| (94) Stago | | 107 | 6,3 | 5,9 | 17 | 0 | | | | | | | 17 | | |
| (149) Siemens (Dade, BN, Dimension) | | 105 | 4,5 | 4,3 | 96 | 0 | | | | | | | 96 | | |
| (175) DIAGON | | 106 | 6,6 | 6,2 | 9 | 0 | | | | | | | 9 | | |
| (177) Mindray | | 106 | 7,7 | 7,3 | 8 | 0 | | | | | | | 8 | | |
| (193) HYPHEN BioMed | | 106 | 6,1 | 5,7 | 8 | 0 | | | | | | | 8 | | |
| Other | | | | | 4 | 0 | | | | | | | 4 | | |
| | | | | | | | | 1x 46, 3x 179 | | | | | | | |
| Sample B | | 55,0 | 4,5 | 8,2 | 171 | | CVP | 55 | 0,85 | 18% | 45,1 | 64,9 | 171 | 162 | 95% |

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of the groups n = 5

EQA round: HKG4/21 - Haemocoagulation Tests

Deadline: 19.11.2021

| Test | [unit] | RoM | SD | CV [%] | N _{tot} | N _{out} | Comparability | | | | | | N _{eva} | N _{suc} | S _{rel} |
|---|--------|------|-------|--------|------------------|------------------|---------------|--|------------------|-----|-------|------|------------------|------------------|------------------|
| | | | | | | | AV | U _{AV} | D _{max} | LL | UL | | | | |
| (177) Antithrombin | | | | | 171 | | | | | | | | 171 | 162 | 95% |
| Samples and groups | [%] | | | | | | | | | | | | | | |
| Sample B | | 55,0 | 4,5 | 8,2 | 171 | | CVP | 55 | 0,85 | 18% | 45,1 | 64,9 | 171 | 162 | 95% |
| (38) IL | | 53,8 | 3,7 | 6,9 | 29 | 0 | | | | | | | | | 29 |
| (94) Stago | | 60,9 | 2,9 | 4,8 | 17 | 0 | | | | | | | | | 17 |
| (149) Siemens (Dade, BN, Dimension) | | 53,6 | 3,4 | 6,3 | 96 | 0 | | | | | | | | | 96 |
| (175) DIAGON | | 58,3 | 4,9 | 8,4 | 9 | 0 | | | | | | | | | 9 |
| (177) Mindray | | 62,7 | 8,1 | 13 | 8 | 0 | | | | | | | | | 8 |
| (193) HYPHEN BioMed | | 56,6 | 2,4 | 4,2 | 8 | 0 | | | | | | | | | 8 |
| Other | | | | | 4 | 0 | | | | | | | | | 4 |
| | | | | | | | | 1x 46, 3x 179 | | | | | | | |
| (172) Prothrombin test (INR) | | | | | 315 | | | | | | | | 306 | 295 | 96% |
| Samples and groups | [-] | | | | | | | | | | | | | | |
| Sample A | | 1,03 | 0,052 | 5,1 | 315 | | CVP | 1,03 | 0,072 | 20% | 0,824 | 1,24 | 315 | 314 | 100% |
| (38) IL | | 1,03 | 0,061 | 5,9 | 35 | 0 | | | | | | | | | 35 |
| (94) Stago | | 1,01 | 0,031 | 3,0 | 38 | 0 | | | | | | | | | 38 |
| (149) Siemens (Dade, BN, Dimension) | | 1,03 | 0,053 | 5,1 | 203 | 0 | | | | | | | | | 203 |
| (175) DIAGON | | 1,05 | 0,062 | 5,9 | 15 | 0 | | | | | | | | | 15 |
| (177) Mindray | | 1,06 | 0,037 | 3,5 | 9 | 0 | | | | | | | | | 9 |
| Other | | | | | 15 | 0 | | | | | | | | | 15 |
| | | | | | | | | 3x 46, 2x 51, 1x 64, 1x 166, 4x 179, 4x 999 | | | | | | | |
| Sample B | | 2,94 | 0,30 | 10 | 315 | | | | | | | | 306 | 295 | 96% |
| (38) IL | | 2,63 | 0,13 | 5,0 | 35 | 0 | CVP | 2,96 | 0,040 | 20% | 2,36 | 3,56 | | | 35 |
| (94) Stago | | 3,04 | 0,19 | 6,3 | 38 | 0 | CVP | 2,96 | 0,040 | 20% | 2,36 | 3,56 | | | 38 |
| (149) Siemens (Dade, BN, Dimension) | | 3,01 | 0,25 | 8,3 | 203 | 0 | CVP | 2,96 | 0,040 | 20% | 2,36 | 3,56 | | | 203 |
| (175) DIAGON | | 2,85 | 0,61 | 21 | 15 | 0 | CVP | 2,96 | 0,040 | 20% | 2,36 | 3,56 | | | 15 |
| (177) Mindray | | 1,92 | 0,082 | 4,3 | 9 | 0 | CVPG | 1,92 | 0,10 | 20% | 1,53 | 2,31 | | | 0 |
| Other | | | | | 15 | 0 | | | | | | | | | 15 |
| | | | | | | | | 3x 46, 2x 51, 1x 64, 1x 166, 4x 179, 4x 999 | | | | | | | |
| (179) Prothrombin test (ratio) | | | | | 313 | | | | | | | | 304 | 291 | 96% |
| Samples and groups | [-] | | | | | | | | | | | | | | |
| Sample A | | 1,03 | 0,051 | 4,9 | 313 | | CVP | 1,03 | 0,070 | 20% | 0,824 | 1,24 | 313 | 312 | 100% |
| (38) IL | | 1,03 | 0,059 | 5,7 | 35 | 0 | | | | | | | | | 35 |
| (94) Stago | | 1,01 | 0,032 | 3,1 | 36 | 0 | | | | | | | | | 36 |
| (149) Siemens (Dade, BN, Dimension) | | 1,03 | 0,051 | 4,9 | 204 | 0 | | | | | | | | | 204 |
| (175) DIAGON | | 1,05 | 0,072 | 6,9 | 14 | 0 | | | | | | | | | 14 |
| (177) Mindray | | 1,06 | 0,037 | 3,5 | 9 | 0 | | | | | | | | | 9 |
| Other | | | | | 15 | 0 | | | | | | | | | 15 |
| | | | | | | | | 1x 27, 3x 46, 2x 51, 1x 64, 1x 166, 3x 179, 4x 999 | | | | | | | |
| Sample B | | 2,86 | 0,27 | 9,5 | 313 | | | | | | | | 304 | 292 | 96% |
| (38) IL | | 2,65 | 0,14 | 5,2 | 35 | 0 | CVP | 2,87 | 0,036 | 20% | 2,29 | 3,45 | | | 35 |
| (94) Stago | | 2,85 | 0,17 | 6,1 | 36 | 0 | CVP | 2,87 | 0,036 | 20% | 2,29 | 3,45 | | | 36 |
| (149) Siemens (Dade, BN, Dimension) | | 2,95 | 0,24 | 8,0 | 204 | 0 | CVP | 2,87 | 0,036 | 20% | 2,29 | 3,45 | | | 204 |
| (175) DIAGON | | 2,53 | 0,42 | 16 | 14 | 0 | CVP | 2,87 | 0,036 | 20% | 2,29 | 3,45 | | | 14 |
| (177) Mindray | | 1,95 | 0,089 | 4,6 | 9 | 0 | CVPG | 1,95 | 0,11 | 20% | 1,56 | 2,34 | | | 0 |
| Other | | | | | 15 | 0 | | | | | | | | | 15 |
| | | | | | | | | 1x 27, 3x 46, 2x 51, 1x 64, 1x 166, 3x 179, 4x 999 | | | | | | | |
| (173) Thrombin time (time) | | | | | 153 | | | | | | | | 134 | 128 | 96% |
| Samples and groups | [s] | | | | | | | | | | | | | | |
| Sample A | | 16,9 | 1,9 | 11 | 153 | | | | | | | | 134 | 131 | 98% |
| (3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago | | 18,1 | 0,81 | 4,5 | 16 | 0 | CVPG | 18,1 | 0,50 | 20% | 14,4 | 21,8 | | | 16 |
| (3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 20,4 | 1,4 | 6,9 | 6 | 0 | CVPG | 20,4 | 2,0 | 20% | 16,3 | 24,5 | | | 6 |
| (5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 15,8 | 1,2 | 7,7 | 80 | 0 | CVPG | 15,8 | 0,33 | 20% | 12,6 | 19 | | | 80 |
| (6) Thrombin units: 3 IU/mL (NIH); (38) IL | | 17,9 | 1,0 | 5,7 | 20 | 0 | CVPG | 17,9 | 0,56 | 20% | 14,3 | 21,5 | | | 20 |
| (7) Thrombin units: 5 IU/mL (NIH); (175) DIAGON | | 20,7 | 3,9 | 19 | 7 | 0 | CVPG | 20,7 | 3,8 | 20% | 16,5 | 24,9 | | | 7 |
| (8) Thrombin units: 8 IU/mL (NIH); (177) Mindray | | 17,1 | 0,89 | 5,2 | 5 | 0 | CVPG | 17,1 | 2,5 | 20% | 13,6 | 20,6 | | | 5 |
| Other | | | | | 19 | 0 | | | | | | | | | 0 |
| | | | | | | | | 1x 1/38, 1x 1/126, 4x 1/149, 2x 1/175, 1x 1/179, 1x 2/149, 1x 2/193, 4x 5/179, 1x 6/998, 2x 7/38, 1x 99/46 | | | | | | | |
| Sample B | | 20,6 | 5,3 | 26 | 153 | | | | | | | | 134 | 130 | 97% |
| (3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago | | 23,5 | 0,75 | 3,2 | 16 | 0 | CVPG | 23,5 | 0,46 | 20% | 18,8 | 28,2 | | | 16 |
| (3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 22,9 | 1,6 | 6,8 | 6 | 0 | CVPG | 22,9 | 2,2 | 20% | 18,3 | 27,5 | | | 6 |
| (5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 17,1 | 1,0 | 6,1 | 80 | 0 | CVPG | 17,1 | 0,29 | 20% | 13,6 | 20,6 | | | 80 |
| (6) Thrombin units: 3 IU/mL (NIH); (38) IL | | 30,1 | 1,5 | 5,0 | 20 | 0 | CVPG | 30,1 | 0,82 | 20% | 24 | 36,2 | | | 20 |
| (7) Thrombin units: 5 IU/mL (NIH); (175) DIAGON | | 26,5 | 2,2 | 8,4 | 7 | 0 | CVPG | 26,5 | 2,2 | 20% | 21,2 | 31,8 | | | 7 |
| (8) Thrombin units: 8 IU/mL (NIH); (177) Mindray | | 22,8 | 1,0 | 4,4 | 5 | 0 | CVPG | 22,8 | 2,8 | 20% | 18,2 | 27,4 | | | 5 |

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of the groups n = 5

EQA round: HKG4/21 - Haemocoagulation Tests

Deadline: 19.11.2021

| Test | [unit] | RoM | SD | CV [%] | N _{tot} | N _{out} | Comparability | | | | | | N _{eva} | N _{suc} | S _{rel} |
|--|--------|-------|-------|--------|------------------|------------------|---------------|-----------------|------------------|-----|-------|------|------------------|------------------|--|
| | | | | | | | AV | U _{AV} | D _{max} | LL | UL | | | | |
| (173) Thrombin time (time) | [s] | | | | 153 | | | | | | | | 134 | 128 | 96% |
| Samples and groups | | | | | | | | | | | | | | | |
| Sample B | | 20,6 | 5,3 | 26 | 153 | | | | | | | | 134 | 130 | 97% |
| Other | | | | | 19 | 0 | | | | | | | 0 | | |
| | | | | | | | | | | | | | | | 1x 1/38, 1x 1/126, 4x 1/149, 2x 1/175, 1x 1/179, 1x 2/149, 1x 2/193, 4x 5/179, 1x 6/998, 2x 7/38, 1x 99/46 |
| (174) Thrombin time (ratio) | [-] | | | | 135 | | | | | | | | 119 | 115 | 97% |
| Samples and groups | | | | | | | | | | | | | | | |
| Sample A | | 1,04 | 0,10 | 10 | 135 | | | | | | | | 119 | 117 | 98% |
| (3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago | | 1,01 | 0,056 | 5,5 | 15 | 0 | CVPG | 1,01 | 0,035 | 20% | 0,808 | 1,22 | 15 | | |
| (3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 1,09 | 0,13 | 12 | 5 | 0 | CVPG | 1,09 | 0,36 | 20% | 0,872 | 1,31 | 5 | | |
| (5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 0,993 | 0,072 | 7,3 | 69 | 0 | CVPG | 0,993 | 0,021 | 20% | 0,794 | 1,2 | 69 | | |
| (6) Thrombin units: 3 IU/mL (NIH); (38) IL | | 1,21 | 0,10 | 8,7 | 18 | 0 | CVPG | 1,21 | 0,060 | 20% | 0,968 | 1,46 | 18 | | |
| (7) Thrombin units: 5 IU/mL (NIH); (175) DIAGON | | 1,07 | 0,089 | 8,3 | 7 | 0 | CVPG | 1,07 | 0,086 | 20% | 0,856 | 1,29 | 7 | | |
| (8) Thrombin units: 8 IU/mL (NIH); (177) Mindray | | 1,07 | 0,11 | 10 | 5 | 0 | CVPG | 1,07 | 0,31 | 20% | 0,856 | 1,29 | 5 | | |
| Other | | | | | 16 | 0 | | | | | | | 0 | | |
| | | | | | | | | | | | | | | | 1x 1/38, 1x 1/126, 2x 1/149, 2x 1/175, 1x 2/94, 1x 2/149, 1x 2/193, 2x 5/179, 1x 6/51, 1x 6/998, 2x 7/38, 1x 99/46 |
| Sample B | | 1,24 | 0,26 | 21 | 135 | | | | | | | | 119 | 115 | 97% |
| (3) Thrombin units: 1,5 IU/mL (NIH); (94) Stago | | 1,32 | 0,071 | 5,3 | 15 | 0 | CVPG | 1,32 | 0,045 | 20% | 1,05 | 1,59 | 15 | | |
| (3) Thrombin units: 1,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 1,32 | 0,17 | 13 | 5 | 0 | CVPG | 1,32 | 0,48 | 20% | 1,05 | 1,59 | 5 | | |
| (5) Thrombin units: 2,5 IU/mL (NIH); (149) Siemens (Dade, BN, Dimension) | | 1,07 | 0,073 | 6,8 | 69 | 0 | CVPG | 1,07 | 0,021 | 20% | 0,856 | 1,29 | 69 | | |
| (6) Thrombin units: 3 IU/mL (NIH); (38) IL | | 2,03 | 0,18 | 8,9 | 18 | 0 | CVPG | 2,03 | 0,10 | 20% | 1,62 | 2,44 | 18 | | |
| (7) Thrombin units: 5 IU/mL (NIH); (175) DIAGON | | 1,35 | 0,074 | 5,5 | 7 | 0 | CVPG | 1,35 | 0,072 | 20% | 1,08 | 1,62 | 7 | | |
| (8) Thrombin units: 8 IU/mL (NIH); (177) Mindray | | 1,41 | 0,15 | 11 | 5 | 0 | CVPG | 1,41 | 0,42 | 20% | 1,12 | 1,7 | 5 | | |
| Other | | | | | 16 | 0 | | | | | | | 0 | | |
| | | | | | | | | | | | | | | | 1x 1/38, 1x 1/126, 2x 1/149, 2x 1/175, 1x 2/94, 1x 2/149, 1x 2/193, 2x 5/179, 1x 6/51, 1x 6/998, 2x 7/38, 1x 99/46 |

st_kn_p

End of report

Printed: 01.12.2021