

## Summary statistics - qualitative results

(Groups: manufacturer of kit)

## EQA round: IDDM2/21 - Type 1 Autoimmune Diabetes

Deadline: 22.10.2021

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

## Test

Sample	Group	N	Result	N <sub>rel</sub>
<b>(472) anti-GAD</b>				
A	(36) Immunotech	>>>	1 Negative	2,9 %
	(49) BioVendor	>>>	3 Negative	8,6 %
	(52) Dynex	>>>	1 Negative	2,9 %
	(66) DRG	>>>	6 Negative	17 %
	(87) Euroimmun	>>>	14 Negative	40 %
	(146) MEDIPAN Diagnostica	>>>	3 Negative	8,6 %
	(999) another manufacturer	>>>	7 Negative	20 %
----- Total per sample: n = 35		-----	Success: 100 %	-----
B	(36) Immunotech	>>>	1 Positive	2,9 %
	(49) BioVendor	>>>	3 Positive	8,6 %
	(52) Dynex	>>>	1 Positive	2,9 %
	(66) DRG	>>>	6 Positive	17 %
	(87) Euroimmun	>>>	14 Positive	40 %
	(146) MEDIPAN Diagnostica	>>>	3 Positive	8,6 %
	(999) another manufacturer	>>>	7 Positive	20 %
----- Total per sample: n = 35		-----	Success: 100 %	-----
Total success: 100 %				

**(474) anti-IA-2**

A	(36) Immunotech	>>>	1 Negative	3,2 %
	(49) BioVendor	>>>	3 Negative	9,7 %
	(52) Dynex	>>>	1 Negative	3,2 %
	(66) DRG	>>>	4 Negative	13 %
	(87) Euroimmun	>>>	11 Negative	35 %
			1 Positive	3,2 %
	(146) MEDIPAN Diagnostica	>>>	2 Negative	6,5 %
	(999) another manufacturer	>>>	8 Negative	26 %
----- Total per sample: n = 31		-----	Success: 97 %	-----
B	(36) Immunotech	>>>	1 Positive	3,2 %
	(49) BioVendor	>>>	3 Positive	9,7 %
	(52) Dynex	>>>	1 Positive	3,2 %
	(66) DRG	>>>	4 Positive	13 %
	(87) Euroimmun	>>>	12 Positive	39 %
	(146) MEDIPAN Diagnostica	>>>	2 Positive	6,5 %
	(999) another manufacturer	>>>	8 Positive	26 %
----- Total per sample: n = 31		-----	Success: 100 %	-----
Total success: 97 %				

st\_kl

End of report

Printed: 30.10.2021