

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability								
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}	
Set 1															
(536) Prothrombin test (CoaguChek)					85							85	84	99%	
Samples and groups															
Sample A1		2,22	0,11	4,9	85		CVP	2,22	0,029	20%	1,77	2,67	85	85	100%
All results		2,22	0,11	4,9	85	0							85		
Sample B1		3,64	0,18	5,0	85		CVP	3,64	0,049	20%	2,91	4,37	85	84	99%
All results		3,64	0,18	5,0	85	0							85		
Set 3															
(543) Prothrombin test (microINR)					28								28	19	68%
Samples and groups															
Sample A3		1,93	0,21	11	28		CVP	1,93	0,097	20%	1,54	2,32	28	25	89%
All results		1,93	0,21	11	28	0							28		
Sample B3		2,71	0,37	14	28		CVP	2,71	0,18	20%	2,16	3,26	28	21	75%
All results		2,71	0,37	14	28	2							28		
Set 5															
(552) Prothrombin test (Xprecia)					16								16	14	88%
Samples and groups															
Sample A5		2,62	0,17	6,4	16		CVP	2,62	0,11	20%	2,09	3,15	16	15	94%
All results		2,62	0,17	6,4	16	1							16		
Sample B5		3,17	0,19	5,9	16		CVP	3,17	0,12	20%	2,53	3,81	16	15	94%
All results		3,17	0,19	5,9	16	1							16		

st_kn_p

End of report

Printed: 21.09.2021