

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability			Comparability				
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set 1														
(536) Prothrombin test (CoaguChek)					160						160	159	99%	
Samples and groups	[-]													
Sample A1		1,56	0,059	3,8	160	CVP	1,56	0,011	20%	1,24	1,88	160	159	99%
All results		1,56	0,059	3,8	160	0						160		
Sample B1		2,44	0,13	5,5	160	CVP	2,44	0,026	20%	1,95	2,93	160	160	100%
All results		2,44	0,13	5,5	160	0						160		
st_kn_p						End of report								Printed: 02.11.2021