

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: LMWH2/21 - Low Molecular Weight Heparin

Deadline: 17.09.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability											
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}
(103) Low Molecular Weight Heparin					15						15	15	100%
----- Samples and groups -----	[IU/mL]												
Sample A		0,956	0,067	7,0	15	CVP	0,912	0,020	25%	0,684	1,14	15	15 100%
(193) HYPHEN BioMed		1,00	0,045	4,5	6							6	
Other					9							9	
						2x 38, 4x 149, 3x 179							
Sample B		1,60	0,15	9,3	15	CVP	1,53	0,033	25%	1,14	1,92	15	15 100%
(193) HYPHEN BioMed		1,70	0,067	3,9	6							6	
Other					9							9	
						2x 38, 4x 149, 3x 179							
st_kn_p						End of report							Printed: 28.09.2021