

EQA round: RET2/21 - Reticulocytes (Microscopy)

Deadline: 19.11.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(211) Reticulocytes					62							62	51	82%
Samples and groups	r.10(exp9)/L]													
Sample A		89,8	22	25	62	CVP	89,8	7,0	45%	49,3	131	62	55	89%
All results		89,8	22	25	62	0						62		
Sample B		264	65	24	62	CVP	264	20	45%	145	383	62	56	90%
All results		264	65	24	62	0						62		

st_kn_p

End of report

Printed: 25.11.2021