

EQA round: RF2/21 - Diagnostics of Rheumatoid Arthritis and ASLO

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Set 1

(231) Rheumat. fac. (scr)

A1	All results	2	Negative	1,2 %
		>>> 161	Positive	99 %
Total per sample: n = 163		Success: 99 %		
B1	All results	>>> 160	Negative	98 %
		3	Positive	1,8 %
Total per sample: n = 163		Success: 98 %		
Total success: 98 %				

(233) Rheumat. factor IgG

A1	All results	>>> 47	Positive	100 %
Total per sample: n = 47		Success: 100 %		
B1	All results	>>> 47	Negative	100 %
Total per sample: n = 47		Success: 100 %		
Total success: 100 %				

(234) Rheumat. factor IgA

A1	All results	>>> 46	Positive	100 %
Total per sample: n = 46		Success: 100 %		
B1	All results	>>> 46	Negative	100 %
Total per sample: n = 46		Success: 100 %		
Total success: 100 %				

(235) Rheumat. factor IgM

A1	All results	>>> 46	Positive	100 %
Total per sample: n = 46		Success: 100 %		
B1	All results	>>> 46	Negative	100 %
Total per sample: n = 46		Success: 100 %		
Total success: 100 %				

(277) anti-CCP IgG

A1	All results	>>> 82	Positive	100 %
Total per sample: n = 82		Success: 100 %		
B1	All results	>>> 82	Negative	100 %
Total per sample: n = 82		Success: 100 %		
Total success: 100 %				

(282) anti-MCV

A1	All results	>>> 20	Positive	100 %
Total per sample: n = 20		Success: 100 %		
B1	All results	>>> 20	Negative	100 %
Total per sample: n = 20		Success: 100 %		
Total success: 100 %				