

Cyklus EHK: BM1/26 - Kostní markery

Stop termín: 10.4.2026

Nastavení: skupiny - R (výrobce reagentů); minimální četnost skupin n = 5

| Zkouška Vzorek Skupina | [jednotka] | RoM | SD | CV [%] | N _{tot} | N _{out} | Srovnatelnost | | | | | N _{eva} | N _{suc} | S _{rel} [%] | |
|---|------------|------|----|-----------|------------------|------------------|---------------|-----------------|------------------|-----|------|------------------|------------------|-------------------------|----|
| | | | | | | | AV | U _{AV} | D _{max} | LL | UL | | | | |
| (432) Telopeptid kolagenu β-CTX-I | [ng/L] | | | | 92 | | | | | | | | 92 | 85 | 92 |
| Vzorek A | | 292 | 22 | 7,6 | 92 | | | | | | | | 92 | 90 | 98 |
| (60) Roche | | 296 | 17 | 5,8 | 82 | 0 | CVPG | 296 | 4,6 | 17% | 245 | 347 | 82 | | |
| (73) Snibe | | 99,7 | 13 | 14 | 10 | 0 | CVPG | 99,7 | 10 | 17% | 82,7 | 117 | 10 | | |
| Vzorek B | | 733 | 99 | 14 | 92 | | | | | | | | 92 | 86 | 93 |
| (60) Roche | | 751 | 83 | 11 | 82 | 0 | CVPG | 751 | 23 | 17% | 623 | 879 | 82 | | |
| (73) Snibe | | 258 | 25 | 9,6 | 10 | 0 | CVPG | 258 | 19 | 17% | 214 | 302 | 10 | | |