

EQA round: BM1/26 - Bone Markers

Deadline: 10.4.2026

Setup: groups - R (manufacturer of kit); minimal size of the groups n = 5

Test Sample Group	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability					N _{eva}	N _{suc}	S _{rel} [%]
							AV	U _{AV}	D _{max}	LL	UL			
(432) Collagen telopeptide β-CTX-I	[ng/L]				92							92	85	92
Sample A		292	22	7,6	92							92	90	98
(60) Roche		296	17	5,8	82	0	CVPG	296	4,6	17%	245	347	82	
(73) Snibe		99,7	13	14	10	0	CVPG	99,7	10	17%	82,7	117	10	
Sample B		733	99	14	92							92	86	93
(60) Roche		751	83	11	82	0	CVPG	751	23	17%	623	879	82	
(73) Snibe		258	25	9,6	10	0	CVPG	258	19	17%	214	302	10	