

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

## EQA round: AT1/21 - Autoantibodies in Thyroid Diseases

Deadline: 19.04.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>Set 2</b>														
<b>(590) Anti-TSH receptor</b>					10						10	8	80%	
Samples and groups	[U/L]													
<b>Sample A2</b>		0,670	0,34	51	10	CVP	0,66	0,065	100%	0	1,32	10	9	90%
(60) Roche		0,800	0,000	0,00	7	0						7		
Other					3	0						3		
							2x 24, 1x 29							
<b>Sample B2</b>		4,33	0,93	22	10	CVP	4,24	0,13	33%	2,84	5,64	10	8	80%
(60) Roche		4,05	0,79	19	7	0						7		
Other					3	0						3		
							2x 24, 1x 29							

st\_kn\_p

End of report

Printed: 23.04.2021