

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Key: >>> ... marks correct (expected) result

N ... number of the results

> ... marks conditionally correct (acceptable) result

Nrel ... relative number of the results

± ... marks the result not evaluated

Test

Sample	Group	N	Result	N _{rel}	
(261) Glucose					
A	(5) Dialab	>>>	1 11,1 - 28 mmol/L	0,33 %	
	(9) Bayer	>>>	1 11,1 - 28 mmol/L	0,33 %	
	(12) Beckman Coulter	>	1 2,81 - 11 mmol/L	0,33 %	
		>>>	4 11,1 - 28 mmol/L	1,3 %	
		>>>	3 > 28 mmol/L	0,98 %	
	(18) Cormay	>	1 2,81 - 11 mmol/L	0,33 %	
		>>>	1 11,1 - 28 mmol/L	0,33 %	
		>>>	1 > 28 mmol/L	0,33 %	
	(46) Erba Lachema		1 Negative, normal, <= 2,8 mmol/L	0,33 %	
		>>>	13 11,1 - 28 mmol/L	4,3 %	
		>>>	29 > 28 mmol/L	9,5 %	
	(49) BioVendor	>>>	3 > 28 mmol/L	0,98 %	
	(60) Roche	>>>	1 11,1 - 28 mmol/L	0,33 %	
		>>>	31 > 28 mmol/L	10 %	
	(63) Sysmex	>>>	2 11,1 - 28 mmol/L	0,66 %	
		>>>	4 > 28 mmol/L	1,3 %	
	(109) Axis-Shield		1 Negative, normal, <= 2,8 mmol/L	0,33 %	
	(125) DOT Diagnostics	>>>	1 > 28 mmol/L	0,33 %	
	(150) Arkray	>>>	25 11,1 - 28 mmol/L	8,2 %	
		>>>	23 > 28 mmol/L	7,5 %	
	(161) YD Diagnostics	>>>	1 11,1 - 28 mmol/L	0,33 %	
		>>>	3 > 28 mmol/L	0,98 %	
	(179) Siemens		1 Negative, normal, <= 2,8 mmol/L	0,33 %	
		>	4 2,81 - 11 mmol/L	1,3 %	
		>>>	26 11,1 - 28 mmol/L	8,5 %	
		>>>	17 > 28 mmol/L	5,6 %	
	(197) Analyticon	>>>	2 > 28 mmol/L	0,66 %	
	(201) Iris Diagnostics		>	2 2,81 - 11 mmol/L	0,66 %
		>>>	8 11,1 - 28 mmol/L	2,6 %	
		>>>	6 > 28 mmol/L	2,0 %	
	(202) Dirui		1 Negative, normal, <= 2,8 mmol/L	0,33 %	
		>	3 2,81 - 11 mmol/L	0,98 %	
		>>>	48 11,1 - 28 mmol/L	16 %	
		>>>	11 > 28 mmol/L	3,6 %	
	(207) 77 Elektronika Kft.	>>>	1 11,1 - 28 mmol/L	0,33 %	
		>>>	21 > 28 mmol/L	6,9 %	
	(999) another manufacturer	>>>	2 11,1 - 28 mmol/L	0,66 %	
		>>>	1 > 28 mmol/L	0,33 %	
<hr/> Total per sample: n = 305 —— Success: 99 %					
B	(5) Dialab	>>>	1 Negative, normal, <= 2,8 mmol/L	0,33 %	
	(9) Bayer	>>>	1 Negative, normal, <= 2,8 mmol/L	0,33 %	
	(12) Beckman Coulter	>>>	8 Negative, normal, <= 2,8 mmol/L	2,6 %	
	(18) Cormay	>>>	2 Negative, normal, <= 2,8 mmol/L	0,66 %	
		>>>	1 2,81 - 11 mmol/L	0,33 %	
	(46) Erba Lachema	>>>	42 Negative, normal, <= 2,8 mmol/L	14 %	
		>>>	1 > 28 mmol/L	0,33 %	
	(49) BioVendor	>>>	3 Negative, normal, <= 2,8 mmol/L	0,98 %	
	(60) Roche	>>>	32 Negative, normal, <= 2,8 mmol/L	10 %	
	(63) Sysmex	>>>	6 Negative, normal, <= 2,8 mmol/L	2,0 %	
	(109) Axis-Shield		1 > 28 mmol/L	0,33 %	
	(125) DOT Diagnostics	>>>	1 Negative, normal, <= 2,8 mmol/L	0,33 %	
	(150) Arkray	>>>	48 Negative, normal, <= 2,8 mmol/L	16 %	
	(161) YD Diagnostics	>>>	4 Negative, normal, <= 2,8 mmol/L	1,3 %	
	(179) Siemens	>>>	47 Negative, normal, <= 2,8 mmol/L	15 %	
		>>>	1 > 28 mmol/L	0,33 %	
	(197) Analyticon	>>>	2 Negative, normal, <= 2,8 mmol/L	0,66 %	
	(201) Iris Diagnostics	>>>	16 Negative, normal, <= 2,8 mmol/L	5,2 %	
	(202) Dirui	>>>	62 Negative, normal, <= 2,8 mmol/L	20 %	
		>>>	1 11,1 - 28 mmol/L	0,33 %	
	(207) 77 Elektronika Kft.	>>>	22 Negative, normal, <= 2,8 mmol/L	7,2 %	
	(999) another manufacturer	>>>	3 Negative, normal, <= 2,8 mmol/L	0,98 %	
<hr/> Total per sample: n = 305 —— Success: 98 %					

Total success: 98 %

(263) Ketone bodies

A	(5) Dialab	>	1 1+, 2+	0,33 %
---	------------	---	----------	--------

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
A	(9) Bayer	>>>	1 3+, 4+	0,33 %
	(12) Beckman Coulter	>	1 1+, 2+	0,33 %
		>>>	7 3+, 4+	2,3 %
	(18) Cormay	>>>	3 3+, 4+	0,98 %
	(46) Erba Lachema		1 Negative, <= 0,5 mmol/L	0,33 %
		>	27 1+, 2+	8,8 %
		>>>	16 3+, 4+	5,2 %
	(49) BioVendor	>	2 1+, 2+	0,65 %
	(60) Roche	>	3 1+, 2+	0,98 %
		>>>	28 3+, 4+	9,2 %
	(63) Sysmex	>	6 1+, 2+	2,0 %
	(109) Axis-Shield		1 Negative, <= 0,5 mmol/L	0,33 %
	(125) DOT Diagnostics	>>>	1 3+, 4+	0,33 %
	(150) Arkray	>>>	48 3+, 4+	16 %
	(161) YD Diagnostics	>	2 1+, 2+	0,65 %
		>>>	2 3+, 4+	0,65 %
	(162) Siemens (Atellica)	>	1 1+, 2+	0,33 %
	(179) Siemens		2 Negative, <= 0,5 mmol/L	0,65 %
		>	28 1+, 2+	9,2 %
		>>>	18 3+, 4+	5,9 %
	(197) Analyticon	>	2 1+, 2+	0,65 %
	(201) Iris Diagnostics	>	5 1+, 2+	1,6 %
		>>>	11 3+, 4+	3,6 %
	(202) Dirui		1 Negative, <= 0,5 mmol/L	0,33 %
		>	20 1+, 2+	6,5 %
		>>>	42 3+, 4+	14 %
	(207) 77 Elektronika Kft.	>	14 1+, 2+	4,6 %
		>>>	8 3+, 4+	2,6 %
	(999) another manufacturer	>	2 1+, 2+	0,65 %
		>>>	1 3+, 4+	0,33 %
Total per sample: n = 306 —— Success: 98 %				
B	(5) Dialab	>>>	1 Negative, <= 0,5 mmol/L	0,33 %
	(9) Bayer	>>>	1 Negative, <= 0,5 mmol/L	0,33 %
	(12) Beckman Coulter	>>>	8 Negative, <= 0,5 mmol/L	2,6 %
	(18) Cormay	>>>	2 Negative, <= 0,5 mmol/L	0,65 %
			1 1+, 2+	0,33 %
	(46) Erba Lachema	>>>	43 Negative, <= 0,5 mmol/L	14 %
			1 1+, 2+	0,33 %
	(49) BioVendor	>>>	3 Negative, <= 0,5 mmol/L	0,98 %
	(60) Roche	>>>	31 Negative, <= 0,5 mmol/L	10 %
	(63) Sysmex	>>>	6 Negative, <= 0,5 mmol/L	2,0 %
	(109) Axis-Shield		1 1+, 2+	0,33 %
	(125) DOT Diagnostics	>>>	1 Negative, <= 0,5 mmol/L	0,33 %
	(150) Arkray	>>>	48 Negative, <= 0,5 mmol/L	16 %
	(161) YD Diagnostics	>>>	4 Negative, <= 0,5 mmol/L	1,3 %
	(162) Siemens (Atellica)	>>>	1 Negative, <= 0,5 mmol/L	0,33 %
	(179) Siemens	>>>	47 Negative, <= 0,5 mmol/L	15 %
			1 3+, 4+	0,33 %
	(197) Analyticon	>>>	2 Negative, <= 0,5 mmol/L	0,65 %
	(201) Iris Diagnostics	>>>	16 Negative, <= 0,5 mmol/L	5,2 %
	(202) Dirui	>>>	62 Negative, <= 0,5 mmol/L	20 %
			1 1+, 2+	0,33 %
	(207) 77 Elektronika Kft.	>>>	22 Negative, <= 0,5 mmol/L	7,2 %
	(999) another manufacturer	>>>	3 Negative, <= 0,5 mmol/L	0,98 %
Total per sample: n = 306 —— Success: 98 %				

Total success: 98 %

(264) Total protein

A	(5) Dialab	>>>	1 > 2,5 g/L	0,33 %
	(9) Bayer	>>>	1 > 2,5 g/L	0,33 %
	(12) Beckman Coulter	>	4 0,51 - 2,5 g/L	1,3 %
		>>>	4 > 2,5 g/L	1,3 %
	(18) Cormay	>>>	3 > 2,5 g/L	0,98 %
	(46) Erba Lachema		1 Negative, < 0,3 g/L	0,33 %
		>	14 0,51 - 2,5 g/L	4,6 %
		>>>	29 > 2,5 g/L	9,5 %
	(49) BioVendor		1 0,3 - 0,5 g/L	0,33 %
	(60) Roche	>	2 0,51 - 2,5 g/L	0,66 %
		>	1 0,51 - 2,5 g/L	0,33 %

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
A	(60) Roche	>>>	30 > 2,5 g/L	9,8 %
	(63) Sysmex	>	2 0,51 - 2,5 g/L	0,66 %
	(109) Axis-Shield	>>>	4 > 2,5 g/L	1,3 %
	(125) DOT Diagnostics		1 Negative, < 0,3 g/L	0,33 %
	(150) Arkray	>	1 0,51 - 2,5 g/L	0,33 %
	(161) YD Diagnostics	>	3 0,51 - 2,5 g/L	0,98 %
	(162) Siemens (Atellica)	>>>	45 > 2,5 g/L	15 %
	(179) Siemens	>>>	4 > 2,5 g/L	1,3 %
	(197) Analyticon	>	1 0,51 - 2,5 g/L	0,33 %
	(201) Iris Diagnostics	>>>	1 > 2,5 g/L	0,33 %
	(202) Dirui	>	8 0,51 - 2,5 g/L	2,6 %
	(207) 77 Elektronika Kft.	>>>	8 > 2,5 g/L	2,6 %
	(999) another manufacturer	>	1 Negative, < 0,3 g/L	0,33 %
		>	10 0,51 - 2,5 g/L	3,3 %
		>>>	52 > 2,5 g/L	17 %
			2 Negative, < 0,3 g/L	0,66 %
			7 0,3 - 0,5 g/L	2,3 %
			> 6 0,51 - 2,5 g/L	2,0 %
		>>>	7 > 2,5 g/L	2,3 %
		>	3 0,51 - 2,5 g/L	0,98 %
Total per sample: n = 305		-----	Success: 95 %	-----
B	(5) Dialab	>>>	1 Negative, < 0,3 g/L	0,33 %
	(9) Bayer	>>>	1 Negative, < 0,3 g/L	0,33 %
	(12) Beckman Coulter	>>>	8 Negative, < 0,3 g/L	2,6 %
	(18) Cormay	>>>	3 Negative, < 0,3 g/L	0,98 %
	(46) Erba Lachema	>>>	43 Negative, < 0,3 g/L	14 %
	(49) BioVendor		1 > 2,5 g/L	0,33 %
	(60) Roche	>>>	31 Negative, < 0,3 g/L	10 %
	(63) Sysmex	>>>	6 Negative, < 0,3 g/L	2,0 %
	(109) Axis-Shield		1 > 2,5 g/L	0,33 %
	(125) DOT Diagnostics	>>>	1 Negative, < 0,3 g/L	0,33 %
	(150) Arkray	>>>	48 Negative, < 0,3 g/L	16 %
	(161) YD Diagnostics	>>>	4 Negative, < 0,3 g/L	1,3 %
	(162) Siemens (Atellica)	>>>	1 Negative, < 0,3 g/L	0,33 %
	(179) Siemens	>>>	46 Negative, < 0,3 g/L	15 %
	(197) Analyticon		1 > 2,5 g/L	0,33 %
	(201) Iris Diagnostics	>>>	2 Negative, < 0,3 g/L	0,66 %
	(202) Dirui	>>>	16 Negative, < 0,3 g/L	5,2 %
	(207) 77 Elektronika Kft.	>>>	62 Negative, < 0,3 g/L	20 %
	(999) another manufacturer		1 > 2,5 g/L	0,33 %
		>>>	20 Negative, < 0,3 g/L	6,6 %
			2 > 2,5 g/L	0,66 %
		>>>	3 Negative, < 0,3 g/L	0,98 %
Total per sample: n = 305		-----	Success: 98 %	-----

Total success: 95 %

(266) Urobilinogen

A	(5) Dialab	>>>	1 3+, 4+	0,33 %
	(9) Bayer	>>>	1 3+, 4+	0,33 %
	(12) Beckman Coulter	>	6 1+, 2+	2,0 %
	(18) Cormay	>>>	2 3+, 4+	0,66 %
	(46) Erba Lachema	>>>	3 3+, 4+	0,98 %
	(49) BioVendor		2 Negative, < 17 µmol/L	0,66 %
	(60) Roche	>	12 1+, 2+	3,9 %
	(63) Sysmex	>>>	29 3+, 4+	9,5 %
	(109) Axis-Shield	>>>	3 3+, 4+	0,98 %
	(125) DOT Diagnostics	>	1 1+, 2+	0,33 %
	(150) Arkray	>>>	30 3+, 4+	9,8 %
	(161) YD Diagnostics	>	5 1+, 2+	1,6 %
	(162) Siemens (Atellica)	>>>	1 3+, 4+	0,33 %
			1 Negative, < 17 µmol/L	0,33 %
			>>>	1 3+, 4+
			>>>	48 3+, 4+
			>>>	4 3+, 4+
			>	1 1+, 2+

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}	
A	(179) Siemens	2	Negative, < 17 µmol/L	0,66 %	
	> 18	1+	2+	5,9 %	
	>>> 28	3+	4+	9,2 %	
	>>> 2	3+	4+	0,66 %	
	> 10	1+	2+	3,3 %	
	>>> 6	3+	4+	2,0 %	
	(202) Dirui	1	Negative, < 17 µmol/L	0,33 %	
	> 21	1+	2+	6,9 %	
	>>> 41	3+	4+	13 %	
	(207) 77 Elektronika Kft.	>>> 22	3+	4+	7,2 %
	(999) another manufacturer	>>> 3	3+	4+	0,98 %
	Total per sample: n = 305		Success: 98 %		
B	(5) Dialab	>>> 1	Negative, < 17 µmol/L	0,33 %	
	(9) Bayer	>>> 1	Negative, < 17 µmol/L	0,33 %	
	(12) Beckman Coulter	>>> 8	Negative, < 17 µmol/L	2,6 %	
	(18) Cormay	>>> 3	Negative, < 17 µmol/L	0,98 %	
	(46) Erba Lachema	>>> 42	Negative, < 17 µmol/L	14 %	
		1	1+, 2+	0,33 %	
	(49) BioVendor	>>> 3	Negative, < 17 µmol/L	0,98 %	
	(60) Roche	>>> 31	Negative, < 17 µmol/L	10 %	
	(63) Sysmex	>>> 6	Negative, < 17 µmol/L	2,0 %	
	(109) Axis-Shield	1	3+, 4+	0,33 %	
	(125) DOT Diagnostics	>>> 1	Negative, < 17 µmol/L	0,33 %	
	(150) Arkray	>>> 48	Negative, < 17 µmol/L	16 %	
	(161) YD Diagnostics	>>> 4	Negative, < 17 µmol/L	1,3 %	
	(162) Siemens (Atellica)	>>> 1	Negative, < 17 µmol/L	0,33 %	
	(179) Siemens	>>> 47	Negative, < 17 µmol/L	15 %	
		1	1+, 2+	0,33 %	
	(197) Analyticon	>>> 2	Negative, < 17 µmol/L	0,66 %	
	(201) Iris Diagnostics	>>> 16	Negative, < 17 µmol/L	5,2 %	
	(202) Dirui	>>> 62	Negative, < 17 µmol/L	20 %	
		1	1+, 2+	0,33 %	
	(207) 77 Elektronika Kft.	>>> 22	Negative, < 17 µmol/L	7,2 %	
	(999) another manufacturer	>>> 3	Negative, < 17 µmol/L	0,98 %	
	Total per sample: n = 305		Success: 99 %		

Total success: 98 %

(268) Bilirubin

A	(5) Dialab	>>> 1	3+, 4+	0,33 %
	(9) Bayer	>>> 1	3+, 4+	0,33 %
	(12) Beckman Coulter	1	Negative	0,33 %
	> 6	1+	2+	2,0 %
	>>> 1	3+	4+	0,33 %
	(18) Cormay	>>> 3	3+, 4+	1,0 %
	(46) Erba Lachema	1	Negative	0,33 %
	(49) BioVendor	>>> 39	3+, 4+	13 %
	(60) Roche	>>> 3	3+, 4+	1,0 %
	(63) Sysmex	>>> 31	3+, 4+	10 %
	(109) Axis-Shield	>>> 5	3+, 4+	1,7 %
	(125) DOT Diagnostics	1	Negative	0,33 %
	(150) Arkray	>>> 48	3+, 4+	16 %
	(161) YD Diagnostics	> 1	1+, 2+	0,33 %
	(162) Siemens (Atellica)	>>> 3	3+, 4+	1,0 %
	(179) Siemens	>>> 1	3+, 4+	0,33 %
		1	Negative	0,33 %
	(197) Analyticon	>>> 47	3+, 4+	16 %
	(201) Iris Diagnostics	>>> 2	3+, 4+	0,66 %
		3	Negative	1,0 %
		> 13	1+, 2+	4,3 %
	(202) Dirui	1	Negative	0,33 %
		> 1	1+, 2+	0,33 %
		>>> 61	3+, 4+	20 %
	(207) 77 Elektronika Kft.	>>> 22	3+, 4+	7,3 %
	(999) another manufacturer	>>> 3	3+, 4+	1,0 %
	Total per sample: n = 301		Success: 97 %	
B	(5) Dialab	>>> 1	Negative	0,33 %
	(9) Bayer	>>> 1	Negative	0,33 %
	(12) Beckman Coulter	>>> 8	Negative	2,7 %
	(18) Cormay	>>> 3	Negative	1,0 %

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
B	(46) Erba Lachema	>>>	39 Negative 1 3+, 4+	13 % 0,33 %
	(49) BioVendor	>>>	3 Negative	1,0 %
	(60) Roche	>>>	31 Negative	10 %
	(63) Sysmex	>>>	5 Negative	1,7 %
	(109) Axis-Shield		1 3+, 4+	0,33 %
	(125) DOT Diagnostics	>>>	1 Negative	0,33 %
	(150) Arkray	>>>	48 Negative	16 %
	(161) YD Diagnostics	>>>	4 Negative	1,3 %
	(162) Siemens (Atellica)	>>>	1 Negative	0,33 %
	(179) Siemens	>>>	47 Negative 1 3+, 4+	16 % 0,33 %
	(197) Analyticon	>>>	2 Negative	0,66 %
	(201) Iris Diagnostics	>>>	16 Negative	5,3 %
	(202) Dirui	>>>	62 Negative 1 3+, 4+	21 % 0,33 %
	(207) 77 Elektronika Kft.	>>>	22 Negative	7,3 %
	(999) another manufacturer	>>>	3 Negative	1,0 %
Total per sample: n = 301		—	Success: 99 %	—

Total success: 97 %

(270) WBC

A	(5) Dialab	>>>	1 > 250 leu/ μ L	0,37 %	
	(9) Bayer	>>>	1 51 - 250 leu/ μ L	0,37 %	
	(12) Beckman Coulter		1 Negative	0,37 %	
	(18) Cormay	>>>	7 51 - 250 leu/ μ L	2,6 %	
		>	2 10 - 50 leu/ μ L	0,75 %	
	(46) Erba Lachema	>>>	1 51 - 250 leu/ μ L	0,37 %	
		>>>	11 > 250 leu/ μ L	4,1 %	
	(49) BioVendor	>>>	1 51 - 250 leu/ μ L	0,37 %	
		>>>	2 > 250 leu/ μ L	0,75 %	
	(60) Roche		> 1 10 - 50 leu/ μ L	0,37 %	
		>>>	6 51 - 250 leu/ μ L	2,2 %	
		>>>	24 > 250 leu/ μ L	9,0 %	
	(63) Sysmex		> 3 10 - 50 leu/ μ L	1,1 %	
		>>>	3 51 - 250 leu/ μ L	1,1 %	
	(109) Axis-Shield		1 Negative	0,37 %	
	(125) DOT Diagnostics	>>>	1 > 250 leu/ μ L	0,37 %	
	(150) Arkray	>>>	12 51 - 250 leu/ μ L	4,5 %	
		>>>	36 > 250 leu/ μ L	13 %	
	(161) YD Diagnostics		> 1 10 - 50 leu/ μ L	0,37 %	
		>>>	1 51 - 250 leu/ μ L	0,37 %	
	(162) Siemens (Atellica)		>>>	1 51 - 250 leu/ μ L	0,37 %
	(179) Siemens		1 Negative	0,37 %	
		>	1 10 - 50 leu/ μ L	0,37 %	
		>>>	20 51 - 250 leu/ μ L	7,5 %	
		>>>	24 > 250 leu/ μ L	9,0 %	
	(197) Analyticon		> 1 10 - 50 leu/ μ L	0,37 %	
		>>>	1 51 - 250 leu/ μ L	0,37 %	
	(201) Iris Diagnostics		> 7 10 - 50 leu/ μ L	2,6 %	
		>>>	7 51 - 250 leu/ μ L	2,6 %	
	(202) Dirui		1 Negative	0,37 %	
		>	4 10 - 50 leu/ μ L	1,5 %	
		>>>	37 51 - 250 leu/ μ L	14 %	
		>>>	20 > 250 leu/ μ L	7,5 %	
	(207) 77 Elektronika Kft.		2 Negative	0,75 %	
		>	2 10 - 50 leu/ μ L	0,75 %	
		>>>	8 51 - 250 leu/ μ L	3,0 %	
		>>>	10 > 250 leu/ μ L	3,7 %	
	(999) another manufacturer		> 1 10 - 50 leu/ μ L	0,37 %	
		>>>	2 51 - 250 leu/ μ L	0,75 %	
Total per sample: n = 267		—	Success: 98 %	—	

B	(5) Dialab	>>>	1 Negative	0,37 %
	(9) Bayer	>>>	1 Negative	0,37 %
	(12) Beckman Coulter	>>>	8 Negative	3,0 %
	(18) Cormay	>>>	3 Negative	1,1 %
	(46) Erba Lachema	>>>	12 Negative	4,5 %
	(49) BioVendor	>>>	3 Negative	1,1 %
	(60) Roche	>>>	31 Negative	12 %

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
B	(63) Sysmex	>>>	6 Negative	2,2 %
	(109) Axis-Shield		1 > 250 leu/ μ L	0,37 %
	(125) DOT Diagnostics	>>>	1 Negative	0,37 %
	(150) Arkray	>>>	48 Negative	18 %
	(161) YD Diagnostics	>>>	2 Negative	0,75 %
	(162) Siemens (Atellica)	>>>	1 Negative	0,37 %
	(179) Siemens	>>>	45 Negative	17 %
			1 > 250 leu/ μ L	0,37 %
	(197) Analyticon	>>>	2 Negative	0,75 %
	(201) Iris Diagnostics	>>>	14 Negative	5,2 %
	(202) Dirui	>>>	61 Negative	23 %
			1 > 250 leu/ μ L	0,37 %
	(207) 77 Elektronika Kft.	>>>	22 Negative	8,2 %
	(999) another manufacturer	>>>	3 Negative	1,1 %
Total per sample: n = 267		—	Success: 99 %	—

Total success: 98 %

(272) Blood

A	(5) Dialab	>>>	1 > 50 ery/ μ L	0,33 %
	(9) Bayer	>>>	1 > 50 ery/ μ L	0,33 %
	(12) Beckman Coulter	>>>	8 > 50 ery/ μ L	2,6 %
	(18) Cormay		1 Negative	0,33 %
		>	1 13 - 50 ery/ μ L	0,33 %
		>>>	1 > 50 ery/ μ L	0,33 %
	(46) Erba Lachema		1 Negative	0,33 %
		>	2 13 - 50 ery/ μ L	0,65 %
		>>>	41 > 50 ery/ μ L	13 %
	(49) BioVendor	>>>	3 > 50 ery/ μ L	0,98 %
	(60) Roche	>>>	31 > 50 ery/ μ L	10 %
	(63) Sysmex	>>>	6 > 50 ery/ μ L	2,0 %
	(109) Axis-Shield		1 Negative	0,33 %
	(125) DOT Diagnostics	>>>	1 > 50 ery/ μ L	0,33 %
	(150) Arkray		> 7 13 - 50 ery/ μ L	2,3 %
		>>>	41 > 50 ery/ μ L	13 %
	(161) YD Diagnostics	>>>	4 > 50 ery/ μ L	1,3 %
	(162) Siemens (Atellica)		> 1 13 - 50 ery/ μ L	0,33 %
	(179) Siemens		1 Negative	0,33 %
		>	6 13 - 50 ery/ μ L	2,0 %
		>>>	41 > 50 ery/ μ L	13 %
	(197) Analyticon	>>>	2 > 50 ery/ μ L	0,65 %
	(201) Iris Diagnostics		> 2 13 - 50 ery/ μ L	0,65 %
		>>>	14 > 50 ery/ μ L	4,6 %
	(202) Dirui		1 Negative	0,33 %
		>	4 13 - 50 ery/ μ L	1,3 %
		>>>	58 > 50 ery/ μ L	19 %
	(207) 77 Elektronika Kft.	>>>	22 > 50 ery/ μ L	7,2 %
	(999) another manufacturer		> 2 13 - 50 ery/ μ L	0,65 %
		>>>	1 > 50 ery/ μ L	0,33 %
Total per sample: n = 306		—	Success: 98 %	—

B	(5) Dialab	>>>	1 Negative	0,33 %
	(9) Bayer	>>>	1 Negative	0,33 %
	(12) Beckman Coulter	>>>	7 Negative	2,3 %
			1 5 - 12 ery/ μ L	0,33 %
	(18) Cormay	>>>	2 Negative	0,65 %
			1 5 - 12 ery/ μ L	0,33 %
	(46) Erba Lachema	>>>	43 Negative	14 %
			1 > 50 ery/ μ L	0,33 %
	(49) BioVendor	>>>	3 Negative	0,98 %
	(60) Roche	>>>	31 Negative	10 %
	(63) Sysmex	>>>	6 Negative	2,0 %
	(109) Axis-Shield		1 > 50 ery/ μ L	0,33 %
	(125) DOT Diagnostics	>>>	1 Negative	0,33 %
	(150) Arkray		>>> 48 Negative	16 %
	(161) YD Diagnostics		>>> 4 Negative	1,3 %
	(162) Siemens (Atellica)		>>> 1 Negative	0,33 %
	(179) Siemens		>>> 47 Negative	15 %
			1 > 50 ery/ μ L	0,33 %
	(197) Analyticon	>>>	2 Negative	0,65 %
	(201) Iris Diagnostics		>>> 16 Negative	5,2 %
	(202) Dirui	>>>	62 Negative	20 %

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
B	(202) Dirui	1	> 50 ery/ μ L	0,33 %
	(207) 77 Elektronika Kft.	>>> 22	Negative	7,2 %
	(999) another manufacturer	>>> 3	Negative	0,98 %
	Total per sample: n = 306	Success: 98 %	-	
	Total success: 98 %			

(273) pH

A	(5) Dialab	>	1 7,5	0,33 %
	(9) Bayer	>>>	1 7,0	0,33 %
	(12) Beckman Coulter	>	7 6,5	2,3 %
		>>>	1 7,0	0,33 %
	(18) Cormay	>	1 6,5	0,33 %
		>>>	2 7,0	0,66 %
	(46) Erba Lachema		1 5,0	0,33 %
		>	3 6,5	0,98 %
		>>>	36 7,0	12 %
		>	1 7,5	0,33 %
		>	2 8,0	0,66 %
			1 9,0	0,33 %
	(49) BioVendor	>	3 6,5	0,98 %
	(60) Roche	>>>	27 7,0	8,9 %
		>	4 7,5	1,3 %
	(63) Sysmex	>>>	1 7,0	0,33 %
		>	5 7,5	1,6 %
	(109) Axis-Shield	>	1 6,0	0,33 %
	(125) DOT Diagnostics	>	1 6,5	0,33 %
	(150) Arkray	>>>	2 7,0	0,66 %
		>	44 7,5	14 %
		>	1 8,0	0,33 %
	(161) YD Diagnostics	>	4 6,5	1,3 %
	(162) Siemens (Atellica)	>>>	1 7,0	0,33 %
	(179) Siemens		1 5,0	0,33 %
		>>>	34 7,0	11 %
		>	12 7,5	3,9 %
	(197) Analyticon	>	1 6,0	0,33 %
		>	1 6,5	0,33 %
	(201) Iris Diagnostics	>	13 6,5	4,3 %
		>>>	3 7,0	0,98 %
	(202) Dirui		1 5,0	0,33 %
		>	3 6,0	0,98 %
		>	5 6,5	1,6 %
		>>>	49 7,0	16 %
		>	6 7,5	2,0 %
	(207) 77 Elektronika Kft.	>	6 6,0	2,0 %
		>	10 6,5	3,3 %
		>>>	5 7,0	1,6 %
		>	1 7,5	0,33 %
	(999) another manufacturer	>	1 6,0	0,33 %
		>>>	1 7,0	0,33 %
		>	1 7,5	0,33 %
	Total per sample: n = 305	Success: 99 %	-	
B	(5) Dialab	>	1 6,0	0,33 %
	(9) Bayer	>>>	1 5,0	0,33 %
	(12) Beckman Coulter	>>>	8 5,0	2,6 %
	(18) Cormay	>>>	2 5,0	0,66 %
		>	1 6,0	0,33 %
	(46) Erba Lachema	>>>	43 5,0	14 %
			1 7,0	0,33 %
	(49) BioVendor	>>>	3 5,0	0,98 %
	(60) Roche	>>>	31 5,0	10 %
	(63) Sysmex	>>>	6 5,0	2,0 %
	(109) Axis-Shield		1 7,0	0,33 %
	(125) DOT Diagnostics	>>>	1 5,0	0,33 %
	(150) Arkray	>>>	46 5,0	15 %
		>	1 5,5	0,33 %
	(161) YD Diagnostics	>>>	4 5,0	1,3 %
	(162) Siemens (Atellica)	>>>	1 5,0	0,33 %
	(179) Siemens	>>>	31 5,0	10 %
		>	12 5,5	3,9 %
		>	3 6,0	0,98 %

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
B	(179) Siemens	1	7,0	0,33 %
	(197) Analyticon	>>>	2 5,0	0,66 %
	(201) Iris Diagnostics	>>>	15 5,0	4,9 %
		>	1 6,0	0,33 %
	(202) Dirui	>>>	50 5,0	16 %
		>	12 5,5	3,9 %
		>	1 6,0	0,33 %
		1	7,0	0,33 %
	(207) 77 Elektronika Kft.	>>>	20 5,0	6,6 %
		>	2 5,5	0,66 %
	(999) another manufacturer	>>>	2 5,0	0,66 %
		>	1 6,0	0,33 %
Total per sample: n = 305		Success: 99 %		-

Total success: 98 %

(274) Nitrite

A	(5) Dialab	>>>	1 Positive	0,37 %
	(9) Bayer	>>>	2 Positive	0,74 %
	(12) Beckman Coulter	>>>	8 Positive	3,0 %
	(18) Cormay	>>>	3 Positive	1,1 %
	(46) Erba Lachema	>>>	13 Positive	4,8 %
	(49) BioVendor	>>>	2 Positive	0,74 %
	(60) Roche	>>>	29 Positive	11 %
	(63) Sysmex	>>>	6 Positive	2,2 %
	(109) Axis-Shield		1 Negative	0,37 %
	(125) DOT Diagnostics	>>>	1 Positive	0,37 %
	(150) Arkray	>>>	48 Positive	18 %
	(161) YD Diagnostics	>>>	3 Positive	1,1 %
	(162) Siemens (Atellica)	>>>	1 Positive	0,37 %
	(179) Siemens		1 Negative	0,37 %
		>>>	45 Positive	17 %
	(197) Analyticon	>>>	2 Positive	0,74 %
	(201) Iris Diagnostics	>>>	16 Positive	5,9 %
	(202) Dirui		1 Negative	0,37 %
		>>>	62 Positive	23 %
	(207) 77 Elektronika Kft.	>>>	22 Positive	8,1 %
	(999) another manufacturer	>>>	3 Positive	1,1 %
Total per sample: n = 270		Success: 99 %		-

B	(5) Dialab	>>>	1 Negative	0,37 %
	(9) Bayer	>>>	2 Negative	0,74 %
	(12) Beckman Coulter	>>>	8 Negative	3,0 %
	(18) Cormay	>>>	1 Negative	0,37 %
			2 Positive	0,74 %
	(46) Erba Lachema	>>>	13 Negative	4,8 %
	(49) BioVendor	>>>	2 Negative	0,74 %
	(60) Roche	>>>	29 Negative	11 %
	(63) Sysmex	>>>	6 Negative	2,2 %
	(109) Axis-Shield		1 Positive	0,37 %
	(125) DOT Diagnostics	>>>	1 Negative	0,37 %
	(150) Arkray	>>>	48 Negative	18 %
	(161) YD Diagnostics	>>>	3 Negative	1,1 %
	(162) Siemens (Atellica)	>>>	1 Negative	0,37 %
	(179) Siemens		1 Negative	0,37 %
		>>>	45 Negative	17 %
	(197) Analyticon	>>>	2 Negative	0,74 %
	(201) Iris Diagnostics	>>>	16 Negative	5,9 %
	(202) Dirui		1 Positive	0,37 %
		>>>	62 Negative	23 %
	(207) 77 Elektronika Kft.	>>>	22 Negative	8,1 %
	(999) another manufacturer	>>>	3 Negative	1,1 %
Total per sample: n = 270		Success: 98 %		-

Total success: 98 %

(275) hCG

A	(5) Dialab	>>>	4 Positive	29 %
	(17) Biogema	>>>	3 Positive	21 %
	(32) Human	>>>	2 Positive	14 %
	(115) BioSystems	>>>	1 Positive	7,1 %
	(999) another manufacturer	>>>	4 Positive	29 %
Total per sample: n = 14		Success: 100 %		-

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: DGP1/21 - Urine Strip Tests Analysis

Deadline: 19.02.2021

Test

Sample	Group	N	Result	N _{rel}
B	(5) Dialab	>>>	4 Negative	29 %
	(17) Biogema	>>>	3 Negative	21 %
	(32) Human	>>>	2 Negative	14 %
	(115) BioSystems	>>>	1 Negative	7,1 %
	(999) another manufacturer	>>>	4 Negative	29 %
<hr/> Total per sample: n = 14		Success: 100 %		<hr/>

Total success: 100 %

st_kl

End of report

Printed: 27.02.2021