

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(150) WBC</b>					51							51	47	92%
Samples and groups	r.10(exp9)/L]													
<b>Sample A</b>		4,93	0,22	4,5	51	CVP	4,99	0,032	15%	4,24	5,74	51	47	92%
All results		4,93	0,22	4,5	51	0						51		
<b>Sample B</b>		7,15	0,27	3,8	51	CVP	7,21	0,040	15%	6,12	8,3	51	50	98%
All results		7,15	0,27	3,8	51	0						51		
<b>(151) RBC</b>					51							51	50	98%
Samples and groups	r.10(exp12)/L]													
<b>Sample A</b>		3,81	0,075	2,0	51	CVP	3,81	0,011	7%	3,54	4,08	51	51	100%
All results		3,81	0,075	2,0	51	0						51		
<b>Sample B</b>		3,95	0,069	1,8	51	CVP	3,96	0,012	7%	3,68	4,24	51	50	98%
All results		3,95	0,069	1,8	51	0						51		
<b>(152) Haemoglobin</b>					51							51	50	98%
Samples and groups	[g/L]													
<b>Sample A</b>		123	2,6	2,1	51	CVP	122	0,40	6%	114	130	51	50	98%
All results		123	2,6	2,1	51	0						51		
<b>Sample B</b>		117	2,3	1,9	51	CVP	116	0,38	6%	109	123	51	51	100%
All results		117	2,3	1,9	51	0						51		
<b>(153) HCT</b>					51							51	50	98%
Samples and groups	[-]													
<b>Sample A</b>		0,376	0,012	3,2	51	CVP	0,376	,0016	10%	0,338	0,414	51	51	100%
All results		0,376	0,012	3,2	51	0						51		
<b>Sample B</b>		0,374	0,013	3,5	51	CVP	0,374	,0022	10%	0,336	0,412	51	50	98%
All results		0,374	0,013	3,5	51	0						51		
<b>(154) MCV</b>					51							51	51	100%
Samples and groups	[fL]													
<b>Sample A</b>		98,9	2,3	2,3	51	CVP	98,7	0,36	10%	88,8	109	51	51	100%
All results		98,9	2,3	2,3	51	0						51		
<b>Sample B</b>		94,7	2,6	2,8	51	CVP	94,5	0,46	10%	85	104	51	51	100%
All results		94,7	2,6	2,8	51	0						51		
<b>(155) Platelets</b>					51							51	51	100%
Samples and groups	r.10(exp9)/L]													
<b>Sample A</b>		265	14	5,2	51	CVP	263	2,0	20%	210	316	51	51	100%
All results		265	14	5,2	51	0						51		
<b>Sample B</b>		339	21	6,2	51	CVP	343	3,3	20%	274	412	51	51	100%
All results		339	21	6,2	51	0						51		
<b>(156) RDW</b>					48							46	42	91%
Samples and groups	[%]													
<b>Sample A</b>		13,5	0,52	3,9	48							46	43	93%
(63) Sysmex		13,4	0,40	3,0	22	1	CVPG	13,3	0,053	10%	11,9	14,7	22	
(177) Mindray		13,3	0,41	3,1	19	1	CVPG	13,3	0,24	10%	11,9	14,7	19	
Other					7	0							5	
<b>Sample B</b>		15,6	0,68	4,4	48							46	42	91%
(63) Sysmex		15,4	0,45	2,9	22	1	CVPG	15,3	0,059	10%	13,7	16,9	22	
(177) Mindray		15,7	0,64	4,0	19	1	CVPG	15,7	0,33	10%	14,1	17,3	19	
Other					7	0							5	
<b>(157) MPV</b>					51							49	48	98%
Samples and groups	[fL]													
<b>Sample A</b>		10,5	0,57	5,4	51							49	48	98%
(63) Sysmex		10,8	0,31	2,9	24	0	CVPG	10,7	0,057	18%	8,77	12,7	24	
(177) Mindray		10,4	0,67	6,4	20	0	CVPG	10,4	0,35	18%	8,52	12,3	20	
Other					7	0							5	
<b>Sample B</b>		11,7	0,67	5,7	51							49	48	98%
(63) Sysmex		11,9	0,29	2,5	24	0	CVPG	11,8	0,062	18%	9,67	14	24	
(177) Mindray		11,9	0,77	6,5	20	0	CVPG	11,8	0,39	18%	9,67	14	20	
Other					7	0							5	
<b>(158) PDW [%]</b>					2							2	2	100%
Samples and groups	[%]													
<b>Sample A</b>		17,2	0,20	1,2	2							2	2	100%

Test	[unit]						Comparability						N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL				
<b>(158) PDW [%]</b>					2								2	2	100%
Samples and groups	[%]														
<b>Sample A</b>		17,2	0,20	1,2	2								2	2	100%
Other					2	0							2		
							1x 12, 1x 157								
<b>Sample B</b>		15,9	0,20	1,3	2								2	2	100%
Other					2	0							2		
							1x 12, 1x 157								
<b>(165) PDW [fL]</b>					23								23	22	96%
Samples and groups	[fL]														
<b>Sample A</b>		13,2	0,90	6,8	23								23	22	96%
(63) Sysmex		13,2	0,90	6,8	23	0	CVPG	12,7	0,16	15%	10,7	14,7	23		
<b>Sample B</b>		14,8	0,49	3,3	23								23	23	100%
(63) Sysmex		14,8	0,49	3,3	23	0	CVPG	14,7	0,13	15%	12,4	17	23		
<b>(166) PDW [-]</b>					14								14	14	100%
Samples and groups	[-]														
<b>Sample A</b>		16,2	0,120,76		14								14	14	100%
(177) Mindray		16,2	0,120,76		14	0	CVPG	16,2	0,072	15%	13,7	18,7	14		
<b>Sample B</b>		15,7	0,110,68		14								14	14	100%
(177) Mindray		15,7	0,110,68		14	0	CVPG	15,7	0,064	15%	13,3	18,1	14		
<b>(160) Neutrophils</b>					36								36	32	89%
Samples and groups	f.10(exp9)/L]														
<b>Sample A</b>		2,47	0,22	8,8	36		CVP	2,48	0,029	25%	1,86	3,1	36	32	89%
All results		2,47	0,22	8,8	36	0							36		
<b>Sample B</b>		3,12	0,20	6,4	36		CVP	3,12	0,027	25%	2,34	3,9	36	34	94%
All results		3,12	0,20	6,4	36	1							36		
<b>(161) Lymphocytes</b>					36								36	35	97%
Samples and groups	f.10(exp9)/L]														
<b>Sample A</b>		1,92	0,093	4,8	36		CVP	1,89	0,017	25%	1,41	2,37	36	35	97%
All results		1,92	0,093	4,8	36	0							36		
<b>Sample B</b>		3,57	0,17	4,7	36		CVP	3,59	0,028	25%	2,69	4,49	36	36	100%
All results		3,57	0,17	4,7	36	0							36		
<b>(162) Monocytes</b>					36								36	33	92%
Samples and groups	f.10(exp9)/L]														
<b>Sample A</b>		0,352	0,092	26	36		CVP	0,437	0,012	50%	0,218	0,656	36	34	94%
All results		0,352	0,092	26	36	0							36		
<b>Sample B</b>		0,325	0,073	22	36		CVP	0,38	0,0100	50%	0,19	0,57	36	35	97%
All results		0,325	0,073	22	36	1							36		
<b>(163) Eosinophils</b>					36								36	33	92%
Samples and groups	f.10(exp9)/L]														
<b>Sample A</b>		0,137	0,025	18	36		CVP	0,14	0,0053	50%	0,07	0,21	36	34	94%
All results		0,137	0,025	18	36	0							36		
<b>Sample B</b>		0,105	0,016	15	36		CVP	0,106	0,0026	50%	0,053	0,159	36	33	92%
All results		0,105	0,016	15	36	1							36		
<b>(164) Basophiles</b>					35								35	34	97%
Samples and groups	f.10(exp9)/L]														
<b>Sample A</b>		0,053	0,032	61	35		CVP	0,067	0,0060	asym.	0	0,15	35	34	97%
All results		0,053	0,032	61	35	0							35		
<b>Sample B</b>		0,027	0,02	73	35		CVP	0,038	0,0039	asym.	0	0,216	35	35	100%
All results		0,027	0,02	73	35	0							35		