

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: LMWH1/21 - Low Molecular Weight Heparin

Deadline: 02.04.2021

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(103) Low Molecular Weight Heparin					14						14	14	100%	
----- Samples and groups -----	[IU/mL]													
Sample A		0,247	0,037	15	14	CVP	0,238	,0089	34%	0,157	0,319	14	14	100%
(193) HYPHEN BioMed		0,245	0,052	21	6							6		
Other					8							8		
						2x 38, 3x 149, 3x 179								
Sample B		0,474	0,04	8,3	14	CVP	0,465	0,013	25%	0,348	0,582	14	14	100%
(193) HYPHEN BioMed		0,470	0,045	9,5	6							6		
Other					8							8		
						2x 38, 3x 149, 3x 179								

st_kn_p

End of report

Printed: 15.04.2021