

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
--------	-------	---	--------	------------------

**Set 1****(560) ANA screening (cell nucleus) - prove**

A1	All results	>>>	18	Positive	100 %
Total per sample: n = 18		Success:	100 %		
B1	All results	>>>	18	Negative	100 %
Total per sample: n = 18		Success:	100 %		
Total success: 100 %					

**(572) ANA screening (cell nucleus) - semi-quantity**

A1	All results	±	1	titre 80	11 %
		±	1	titre 160	11 %
		±	2	titre 1280	22 %
		±	5	titre 2560	56 %
Total per sample: n = 9		Success (only evaluated results):	not assessed		
B1	All results	±	1	titre 0 (not measurable)	11 %
		±	5	titre 40	56 %
		±	3	titre 80	33 %
Total per sample: n = 9		Success (only evaluated results):	not assessed		

**(561) ANA - type of fluorescence - nucleus**

A1	All results	>>>	7	Homogenous	50 %
		>>>	5	Homogenous; Coarse or fine speckled	36 %
		>>>	1	Homogenous; DFS 70 like	7,1 %
		>>>	1	Coarse or fine speckled; DFS 70 like	7,1 %
Total per sample: n = 14		Success:	100 %		
B1	All results	>>>	13	No fluorescence	93 %
			1	Coarse or fine speckled	7,1 %
Total per sample: n = 14		Success:	93 %		
Total success: 93 %					

**(563) ANA screening (cytoplasm) - prove**

A1	All results	>>>	14	Negative	100 %
Total per sample: n = 14		Success:	100 %		
B1	All results	>>>	14	Negative	100 %
Total per sample: n = 14		Success:	100 %		
Total success: 100 %					

**(573) ANA screening (cytoplasm) - semi-quantity**

A1	All results	±	3	titre 0 (not measurable)	33 %
		±	4	titre 40	44 %
		±	2	titre 80	22 %
Total per sample: n = 9		Success (only evaluated results):	not assessed		
B1	All results	±	3	titre 0 (not measurable)	33 %
		±	4	titre 40	44 %
		±	2	titre 80	22 %
Total per sample: n = 9		Success (only evaluated results):	not assessed		

**(564) ANA - type of fluorescence - cytoplasm**

A1	All results	>>>	14	No fluorescence	100 %
Total per sample: n = 14		Success:	100 %		
B1	All results	>>>	14	No fluorescence	100 %
Total per sample: n = 14		Success:	100 %		
Total success: 100 %					

**Set 2****(565) ANCA imunofluorescence - prove**

A2	All results	>>>	10	Positive	100 %
Total per sample: n = 10		Success:	100 %		
B2	All results	>>>	10	Negative	100 %
Total per sample: n = 10		Success:	100 %		
Total success: 100 %					

**(574) ANCA IF (ethanol) - semi-quantity**

A2	All results	±	2	titre 40	50 %
		±	1	titre 80	25 %
		±	1	titre 320	25 %
Total per sample: n = 4		Success (only evaluated results):	not assessed		
B2	All results	±	1	titre 0 (not measurable)	25 %
		±	2	titre 20	50 %
		±	1	titre 40	25 %

## Test

Sample	Group	N	Result	N <sub>rel</sub>
Total per sample: n = 4		Success (only evaluated results): not assessed		
<b>(578) ANCA IF (formalin) - semi-quantity</b>				
A2	All results	± 2	titre 40	50 %
		± 2	titre 160	50 %
Total per sample: n = 4		Success (only evaluated results): not assessed		
B2	All results	± 1	titre 0 (not measurable)	25 %
		± 2	titre 20	50 %
		± 1	titre 40	25 %
Total per sample: n = 4		Success (only evaluated results): not assessed		
<b>(566) ANCA - type of fluorescence</b>				
A2	All results	>>> 8	Perinuclear fluorescence	89 %
		>>> 1	Perinuclear fluorescence; ANA	11 %
Total per sample: n = 9		Success: 100 %		
B2	All results	>>> 9	No fluorescence	100 %
Total per sample: n = 9		Success: 100 %		
Total success: 100 %				
<b>(569) anti-myeloperoxidase</b>				
A2	All results	>>> 17	Positive	100 %
Total per sample: n = 17		Success: 100 %		
B2	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 100 %				
<b>(570) anti-proteinase III</b>				
A2	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
B2	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 100 %				
<b>Set 3</b>				
<b>(580) anti-ENA, antigen SS-A/Ro-60</b>				
A3	All results	>>> 16	Positive	100 %
Total per sample: n = 16		Success: 100 %		
B3	All results	>>> 15	Negative	94 %
		1	Inconclusive result (borderline)	6,3 %
Total per sample: n = 16		Success: 94 %		
Total success: 94 %				
<b>(581) anti-ENA, antigen SS-B/La</b>				
A3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
B3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
Total success: 100 %				
<b>(582) anti-ENA, antigen Scl-70</b>				
A3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
B3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
Total success: 100 %				
<b>(583) anti-ENA, antigen RNP</b>				
A3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
B3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
Total success: 100 %				
<b>(584) anti-ENA, antigen Jo-1</b>				
A3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
B3	All results	>>> 16	Negative	100 %
Total per sample: n = 16		Success: 100 %		
Total success: 100 %				
<b>(585) anti-centromere</b>				
A3	All results	>>> 14	Negative	100 %
Total per sample: n = 14		Success: 100 %		
B3	All results	>>> 10	Negative	71 %
		2	Inconclusive result (borderline)	14 %

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
B3	All results	2	Positive	14 %
Total per sample: n = 14		Success: 71 %		
Total success: 71 %				

**Set 4****(588) anti-dsDNA immunofluor. - prove**

A4	All results	3	Negative	33 %
		>>> 6	Positive	67 %
Total per sample: n = 9		Success: 67 %		
B4	All results	>>> 9	Negative	100 %
Total per sample: n = 9		Success: 100 %		
Total success: 67 %				

**(577) anti-dsDNA immunofluor. - semi-quantity**

A4	All results	± 1	titre 10	25 %
		± 2	titre 160	50 %
		± 1	titre 320	25 %
Total per sample: n = 4		Success (only evaluated results): not assessed		
B4	All results	± 4	titre 10	100 %
Total per sample: n = 4		Success (only evaluated results): not assessed		

**(589) anti-dsDNA other methods**

A4	All results	1	Negative	8,3 %
		>>> 11	Positive	92 %
Total per sample: n = 12		Success: 92 %		
B4	All results	>>> 12	Negative	100 %
Total per sample: n = 12		Success: 100 %		
Total success: 92 %				

**Set 5****(591) anti-nukleosomes IgG**

A5	All results	>>> 12	Positive	100 %
Total per sample: n = 12		Success: 100 %		
B5	All results	>>> 12	Negative	100 %
Total per sample: n = 12		Success: 100 %		
Total success: 100 %				