

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

## Test

Sample	Group	N	Result	N <sub>rel</sub>
--------	-------	---	--------	------------------

## Set 1

**(575) Anti-thyreoglobulin**

A1	(0) not specified	>>>	1	Positive	0,53 %
	(1) Abbott	>>>	60	Positive	32 %
	(12) Beckman Coulter	>>>	16	Positive	8,6 %
	(15) bioMérieux	>>>	1	Positive	0,53 %
	(25) TOSOH	>>>	3	Positive	1,6 %
	(29) Siemens (Immulin)	>>>	14	Positive	7,5 %
	(36) Immunotech	>>>	1	Positive	0,53 %
	(60) Roche	>>>	1	Negative	0,53 %
		>>>	60	Positive	32 %
	(87) Euroimmun	>>>	1	Positive	0,53 %
	(135) OrgenTec	>>>	1	Positive	0,53 %
	(162) Siemens (Atellica)	>>>	14	Positive	7,5 %
	(179) Siemens	>>>	13	Positive	7,0 %
	(999) another manufacturer	>>>	1	Positive	0,53 %
----- Total per sample: n = 187 -----		Success: 99 %			
B1	(0) not specified	>>>	1	Negative	0,53 %
	(1) Abbott	>>>	60	Negative	32 %
	(12) Beckman Coulter	>>>	16	Negative	8,6 %
	(15) bioMérieux	>>>	1	Negative	0,53 %
	(25) TOSOH	>>>	3	Negative	1,6 %
	(29) Siemens (Immulin)	>>>	14	Negative	7,5 %
	(36) Immunotech	>>>	1	Negative	0,53 %
	(60) Roche	>>>	61	Negative	33 %
	(87) Euroimmun	>>>	1	Negative	0,53 %
	(135) OrgenTec	>>>	1	Negative	0,53 %
	(162) Siemens (Atellica)	>>>	14	Negative	7,5 %
	(179) Siemens	>>>	13	Negative	7,0 %
	(999) another manufacturer	>>>	1	Negative	0,53 %
----- Total per sample: n = 187 -----		Success: 100 %			
Total success: 99 %					

**(576) Anti-microsomes (TPO)**

A1	(0) not specified	>>>	1	Positive	0,52 %
	(1) Abbott	>>>	61	Positive	32 %
	(12) Beckman Coulter	>>>	16	Positive	8,3 %
	(15) bioMérieux	>>>	1	Positive	0,52 %
	(25) TOSOH	>>>	4	Positive	2,1 %
	(29) Siemens (Immulin)	>>>	15	Positive	7,8 %
	(60) Roche	>>>	62	Positive	32 %
	(87) Euroimmun	>>>	1	Positive	0,52 %
	(135) OrgenTec	>>>	1	Positive	0,52 %
	(162) Siemens (Atellica)	>>>	14	Positive	7,3 %
	(179) Siemens	>>>	15	Positive	7,8 %
	(999) another manufacturer	>>>	1	Positive	0,52 %
----- Total per sample: n = 192 -----		Success: 100 %			
B1	(0) not specified	>>>	1	Negative	0,52 %
	(1) Abbott	>>>	61	Negative	32 %
	(12) Beckman Coulter	>>>	16	Negative	8,3 %
	(15) bioMérieux	>>>	1	Negative	0,52 %
	(25) TOSOH	>>>	4	Negative	2,1 %
	(29) Siemens (Immulin)	>>>	15	Negative	7,8 %
	(60) Roche	>>>	62	Negative	32 %
	(87) Euroimmun	>>>	1	Negative	0,52 %
	(135) OrgenTec	>>>	1	Negative	0,52 %
	(162) Siemens (Atellica)	>>>	14	Negative	7,3 %
	(179) Siemens	>>>	15	Negative	7,8 %
	(999) another manufacturer	>>>	1	Negative	0,52 %
----- Total per sample: n = 192 -----		Success: 100 %			
Total success: 100 %					

Test		N	Result	N <sub>rel</sub>
Sample	Group			
<b>Set 2</b>				
<b>(590) Anti-TSH receptor</b>				
A2	(0) not specified	>>> 1	Positive	1,0 %
	(1) Abbott	>>> 12	Positive	12 %
	(24) B.R.A.H.M.S.	>>> 1	Positive	1,0 %
	(29) Siemens (Immulin)	>>> 10	Positive	10 %
	(36) Immunotech	>>> 1	Positive	1,0 %
	(60) Roche	>>> 69	Positive	71 %
	(87) Euroimmun	>>> 1	Positive	1,0 %
	(999) another manufacturer	>>> 2	Positive	2,1 %
Total per sample: n = 97		Success: 100 %		
B2	(0) not specified	>>> 1	Negative	1,0 %
	(1) Abbott	>>> 11	Negative	11 %
		1	Inconclusive result (borderline)	1,0 %
	(24) B.R.A.H.M.S.	>>> 1	Negative	1,0 %
	(29) Siemens (Immulin)	>>> 10	Negative	10 %
	(36) Immunotech	>>> 1	Negative	1,0 %
	(60) Roche	>>> 69	Negative	71 %
	(87) Euroimmun	>>> 1	Negative	1,0 %
	(999) another manufacturer	>>> 2	Negative	2,1 %
Total per sample: n = 97		Success: 99 %		
Total success: 99 %				