

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

EQA round: CC2/22 - Cystatin C

Deadline: 26.08.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability											
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}
(106) Cystatin C					10						10	10	100%
----- Samples and groups -----	[mg/L]												
Sample A		1,24	0,13	10	10	CVP	1,21	0,031	18%	0,992	1,43	10	10 100%
Other					10							10	
							1x 3, 2x 46, 1x 58, 2x 60, 1x 75, 1x 91, 1x 162, 1x 179						
Sample B		2,74	0,19	7,0	10	CVP	2,7	0,063	18%	2,21	3,19	10	10 100%
Other					10							10	
							1x 3, 2x 46, 1x 58, 2x 60, 1x 75, 1x 91, 1x 162, 1x 179						

st_kn_p

End of report

Printed: 30.08.2022