

EQA round: DGP1/22 - Urine Strip Tests Analysis

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
(268) Bilirubin				
A	All results	>>> 298	Negative	100 %
Total per sample: n = 298		Success: 100 %		
B	All results	1	Not specified	0,34 %
		>>> 295	Negative	99 %
		1	1+, 2+	0,34 %
		1	3+, 4+	0,34 %
Total per sample: n = 298		Success: 99 %		
Total success: 99 %				

(264) Total protein

A	All results	>>> 300	Negative, < 0,3 g/L	100 %
		1	0,51 - 2,5 g/L	0,33 %
Total per sample: n = 301		Success: 100 %		
B	All results	5	Negative, < 0,3 g/L	1,7 %
		>>> 150	0,3 - 0,5 g/L	50 %
		>>> 135	0,51 - 2,5 g/L	45 %
		> 11	> 2,5 g/L	3,7 %
Total per sample: n = 301		Success: 98 %		
Total success: 98 %				

(261) Glucose

A	All results	>>> 299	Negative, normal, <= 2,8 mmol/L	99 %
		2	2,81 - 11 mmol/L	0,66 %
Total per sample: n = 301		Success: 99 %		
B	All results	2	Negative, normal, <= 2,8 mmol/L	0,66 %
		> 32	2,81 - 11 mmol/L	11 %
		>>> 221	11,1 - 28 mmol/L	73 %
		> 46	> 28 mmol/L	15 %
Total per sample: n = 301		Success: 99 %		
Total success: 99 %				

(275) hCG

A	All results	>>> 14	Negative	100 %
Total per sample: n = 14		Success: 100 %		
B	All results	>>> 14	Negative	100 %
Total per sample: n = 14		Success: 100 %		
Total success: 100 %				

(263) Ketone bodies

A	All results	>>> 299	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 299		Success: 100 %		
B	All results	>>> 298	Negative, <= 0,5 mmol/L	100 %
		1	1+, 2+	0,33 %
Total per sample: n = 299		Success: 100 %		
Total success: 100 %				

(272) Blood

A	All results	>>> 298	Negative	99 %
		2	5 - 12 ery/ μ L	0,66 %
		1	> 50 ery/ μ L	0,33 %
Total per sample: n = 301		Success: 99 %		
B	All results	2	Negative	0,66 %
		2	5 - 12 ery/ μ L	0,66 %
		> 22	13 - 50 ery/ μ L	7,3 %
		>>> 275	> 50 ery/ μ L	91 %
Total per sample: n = 301		Success: 99 %		
Total success: 98 %				

(270) WBC

A	All results	>>> 265	Negative	100 %
Total per sample: n = 265		Success: 100 %		
B	All results	>>> 260	Negative	98 %
		5	10 - 50 leu/ μ L	1,9 %
Total per sample: n = 265		Success: 98 %		
Total success: 98 %				

(274) Nitrite

A	All results	>>> 264	Negative	99 %
		2	Positive	0,75 %
Total per sample: n = 266		Success: 99 %		

EQA round: DGP1/22 - Urine Strip Tests Analysis

Test		N	Result	N _{rel}
Sample	Group			
B	All results	>>> 264	Negative	99 %
		2	Positive	0,75 %
Total per sample: n = 266		Success: 99 %		
Total success: 99 %				

(273) pH

A	All results	>>> 219	5,0	73 %
		> 64	5,5	21 %
		> 16	6,0	5,3 %
		2	6,5	0,66 %
Total per sample: n = 301		Success: 99 %		
Total success: 99 %				
B	All results	3	5,0	1,0 %
		> 1	5,5	0,33 %
		> 58	6,0	19 %
		>>> 202	6,5	67 %
		> 37	7,0	12 %
Total per sample: n = 301		Success: 99 %		
Total success: 99 %				

(266) Urobilinogen

A	All results	>>> 300	Negative, < 17 µmol/L	100 %
Total per sample: n = 300		Success: 100 %		
B	All results	>>> 298	Negative, < 17 µmol/L	99 %
		2	1+, 2+	0,67 %
Total per sample: n = 300		Success: 99 %		
Total success: 99 %				

st_kl

End of report

Printed: 24.02.2022