

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(268) Bilirubin</b>				
A	(1) Visual	>>> 30	Negative	10 %
	(2) Reflectance (reader)	>>> 268	Negative	90 %
Total per sample: n = 298		Success: 100 %		
B	(1) Visual	1	Not specified	0,34 %
		>>> 29	Negative	9,7 %
	(2) Reflectance (reader)	>>> 266	Negative	89 %
		1	1+, 2+	0,34 %
		1	3+, 4+	0,34 %
Total per sample: n = 298		Success: 99 %		
Total success: 99 %				

**(264) Total protein**

A	(1) Visual	>>> 32	Negative, < 0,3 g/L	11 %
	(2) Reflectance (reader)	>>> 268	Negative, < 0,3 g/L	89 %
		1	0,51 - 2,5 g/L	0,33 %
Total per sample: n = 301		Success: 100 %		
B	(1) Visual	>>> 19	0,3 - 0,5 g/L	6,3 %
		>>> 12	0,51 - 2,5 g/L	4,0 %
		> 1	> 2,5 g/L	0,33 %
	(2) Reflectance (reader)	5	Negative, < 0,3 g/L	1,7 %
		>>> 131	0,3 - 0,5 g/L	44 %
		>>> 123	0,51 - 2,5 g/L	41 %
		> 10	> 2,5 g/L	3,3 %
Total per sample: n = 301		Success: 98 %		
Total success: 98 %				

**(261) Glucose**

A	(1) Visual	>>> 32	Negative, normal, <= 2,8 mmol/L	11 %
	(2) Reflectance (reader)	>>> 267	Negative, normal, <= 2,8 mmol/L	89 %
		2	2,81 - 11 mmol/L	0,66 %
Total per sample: n = 301		Success: 99 %		
B	(1) Visual	>>> 14	11,1 - 28 mmol/L	4,7 %
		> 18	> 28 mmol/L	6,0 %
	(2) Reflectance (reader)	2	Negative, normal, <= 2,8 mmol/L	0,66 %
		> 32	2,81 - 11 mmol/L	11 %
		>>> 207	11,1 - 28 mmol/L	69 %
		> 28	> 28 mmol/L	9,3 %
Total per sample: n = 301		Success: 99 %		
Total success: 99 %				

**(275) hCG**

A	(1) Visual	>>> 12	Negative	86 %
	(2) Reflectance (reader)	>>> 2	Negative	14 %
Total per sample: n = 14		Success: 100 %		
B	(1) Visual	>>> 12	Negative	86 %
	(2) Reflectance (reader)	>>> 2	Negative	14 %
Total per sample: n = 14		Success: 100 %		
Total success: 100 %				

**(263) Ketone bodies**

A	(1) Visual	>>> 32	Negative, <= 0,5 mmol/L	11 %
	(2) Reflectance (reader)	>>> 267	Negative, <= 0,5 mmol/L	89 %
Total per sample: n = 299		Success: 100 %		
B	(1) Visual	>>> 32	Negative, <= 0,5 mmol/L	11 %
	(2) Reflectance (reader)	>>> 266	Negative, <= 0,5 mmol/L	89 %
		1	1+, 2+	0,33 %
Total per sample: n = 299		Success: 100 %		
Total success: 100 %				

**(272) Blood**

A	(1) Visual	>>> 31	Negative	10 %
		1	5 - 12 ery/ $\mu$ L	0,33 %
	(2) Reflectance (reader)	>>> 267	Negative	89 %
		1	5 - 12 ery/ $\mu$ L	0,33 %
		1	> 50 ery/ $\mu$ L	0,33 %
Total per sample: n = 301		Success: 99 %		
B	(1) Visual	> 3	13 - 50 ery/ $\mu$ L	1,0 %
		>>> 29	> 50 ery/ $\mu$ L	9,6 %

Test	Sample	Group	N	Result	N <sub>rel</sub>	
B	(2) Reflectance (reader)		2	Negative	0,66 %	
			2	5 - 12 ery/ $\mu$ L	0,66 %	
			> 19	13 - 50 ery/ $\mu$ L	6,3 %	
			>>> 246	> 50 ery/ $\mu$ L	82 %	
		Total per sample: n = 301		Success: 99 %		
Total success: 98 %						
<b>(270) WBC</b>						
A	(1) Visual		>>> 6	Negative	2,3 %	
		(2) Reflectance (reader)		>>> 259	Negative	98 %
			Total per sample: n = 265		Success: 100 %	
B	(1) Visual		>>> 4	Negative	1,5 %	
			2	10 - 50 leu/ $\mu$ L	0,75 %	
		(2) Reflectance (reader)		>>> 256	Negative	97 %
				3	10 - 50 leu/ $\mu$ L	1,1 %
		Total per sample: n = 265		Success: 98 %		
Total success: 98 %						
<b>(274) Nitrite</b>						
A	(1) Visual		>>> 8	Negative	3,0 %	
		(2) Reflectance (reader)		>>> 256	Negative	96 %
				2	Positive	0,75 %
Total per sample: n = 266		Success: 99 %				
B	(1) Visual		>>> 8	Negative	3,0 %	
		(2) Reflectance (reader)		>>> 256	Negative	96 %
				2	Positive	0,75 %
Total per sample: n = 266		Success: 99 %				
Total success: 99 %						
<b>(273) pH</b>						
A	(1) Visual		>>> 32	5,0	11 %	
		(2) Reflectance (reader)		>>> 187	5,0	62 %
				> 64	5,5	21 %
				> 16	6,0	5,3 %
				2	6,5	0,66 %
Total per sample: n = 301		Success: 99 %				
B	(1) Visual		2	5,0	0,66 %	
			> 1	5,5	0,33 %	
			> 16	6,0	5,3 %	
			>>> 11	6,5	3,7 %	
			> 2	7,0	0,66 %	
	(2) Reflectance (reader)		1	5,0	0,33 %	
			> 42	6,0	14 %	
			>>> 191	6,5	63 %	
			> 35	7,0	12 %	
		Total per sample: n = 301		Success: 99 %		
Total success: 99 %						
<b>(266) Urobilinogen</b>						
A	(1) Visual		>>> 32	Negative, < 17 $\mu$ mol/L	11 %	
		(2) Reflectance (reader)		>>> 268	Negative, < 17 $\mu$ mol/L	89 %
			Total per sample: n = 300		Success: 100 %	
B	(1) Visual		>>> 31	Negative, < 17 $\mu$ mol/L	10 %	
			1	1+, 2+	0,33 %	
		(2) Reflectance (reader)		>>> 267	Negative, < 17 $\mu$ mol/L	89 %
				1	1+, 2+	0,33 %
		Total per sample: n = 300		Success: 99 %		
Total success: 99 %						