

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
(268) Bilirubin				
A	All results	>>> 26	Negative	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	>>> 26	Negative	100 %
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				
(264) Total protein				
A	All results	>>> 26	Negative, < 0,3 g/L	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	>>> 17	0,3 - 0,5 g/L	65 %
		>>> 9	0,51 - 2,5 g/L	35 %
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				
(261) Glucose				
A	All results	>>> 26	Negative, normal, <= 2,8 mmol/L	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	> 5	2,81 - 11 mmol/L	19 %
		>>> 19	11,1 - 28 mmol/L	73 %
		> 2	> 28 mmol/L	7,7 %
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				
(275) hCG				
A	All results	>>> 5	Negative	100 %
Total per sample: n = 5		Success: 100 %		
B	All results	>>> 5	Negative	100 %
Total per sample: n = 5		Success: 100 %		
Total success: 100 %				
(263) Ketone bodies				
A	All results	>>> 24	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	>>> 24	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(272) Blood				
A	All results	>>> 26	Negative	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	1	5 - 12 ery/ μ L	3,8 %
		> 2	13 - 50 ery/ μ L	7,7 %
		>>> 23	> 50 ery/ μ L	88 %
Total per sample: n = 26		Success: 96 %		
Total success: 96 %				
(270) WBC				
A	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(274) Nitrite				
A	All results	>>> 25	Negative	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	>>> 24	Negative	96 %
		1	Positive	4,0 %
Total per sample: n = 25		Success: 96 %		
Total success: 96 %				
(273) pH				
A	All results	>>> 18	5,0	69 %
		> 5	5,5	19 %
		> 3	6,0	12 %
Total per sample: n = 26		Success: 100 %		
B	All results	> 2	6,0	7,7 %
		>>> 22	6,5	85 %
		> 2	7,0	7,7 %

Test

Sample	Group	N	Result	N _{rel}
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				

(266) Urobilinogen

A	All results	>>> 26	Negative, < 17 µmol/L	100 %
Total per sample: n = 26		Success: 100 %		
B	All results	>>> 25	Negative, < 17 µmol/L	96 %
		1	1+, 2+	3,8 %
Total per sample: n = 26		Success: 96 %		
Total success: 96 %				