

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: FC1/22 - Calprotectin in Stool

Deadline: 04.03.2022

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(223) Calprotectin				
A	(49) BioVendor	>>>	1 Positive	1,4 %
	(87) Euroimmun	>>>	1 Positive	1,4 %
	(135) OrgenTec	>>>	7 Positive	9,7 %
	(164) DiaSorin	>>>	7 Positive	9,7 %
	(166) Thermo Fisher	>>>	6 Positive	8,3 %
	(199) Eurolab Lambda	>>>	1 Positive	1,4 %
	(200) BodiTech	>>>	1 Positive	1,4 %
	(211) INOVA Diagnostics	>>>	5 Positive	6,9 %
	(219) Bühlmann		1 Negative	1,4 %
			3 Inconclusive result (borderline)	4,2 %
		>>>	27 Positive	38 %
	(999) another manufacturer	>>>	12 Positive	17 %
----- Total per sample: n = 72 -----		Success: 94 %	-----	
B	(49) BioVendor	>>>	1 Negative	1,4 %
	(87) Euroimmun	>>>	1 Negative	1,4 %
	(135) OrgenTec	>>>	7 Negative	9,7 %
	(164) DiaSorin	>>>	6 Negative	8,3 %
			1 Positive	1,4 %
	(166) Thermo Fisher	>>>	6 Negative	8,3 %
	(199) Eurolab Lambda	>>>	1 Negative	1,4 %
	(200) BodiTech	>>>	1 Negative	1,4 %
	(211) INOVA Diagnostics	>>>	5 Negative	6,9 %
	(219) Bühlmann	>>>	31 Negative	43 %
	(999) another manufacturer	>>>	12 Negative	17 %
----- Total per sample: n = 72 -----		Success: 99 %	-----	
Total success: 93 %				