

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of the groups n = 5

## EQA round: FC2/22 - Calprotectin in Stool

Deadline: 23.09.2022

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(223) Calprotectin</b>					10							10	9	90%
----- Samples and groups -----	[µg/g]													
<b>Sample A</b>		25,6	9,3	36	10	CVP	40,1	6,5	100%	0	80,2	10	9	90%
Other					10	1						10		
							2x 135, 1x 164, 1x 199, 1x 200, 1x 211, 4x 999							
<b>Sample B</b>		11,6	6,6	57	10	CVP	18,4	3,2	100%	0	36,8	10	9	90%
Other					10	1						10		
							2x 135, 1x 164, 1x 199, 1x 200, 1x 211, 4x 999							

st\_kn\_p

End of report

Printed: 04.10.2022