

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Test					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(247) Gammopathies (quantity 1) (plasma)	[g/L]				88							88	67	76%
Samples and groups														
Sample A		20,6	3,0	15	88	CVP	20,6	0,79	33%	13,8	27,4	88	83	94%
All results		20,6	3,0	15	88	0						88		
Sample B		7,64	1,8	23	88	CVP	7,64	0,47	33%	5,11	10,2	88	70	80%
All results		7,64	1,8	23	88	2						88		
(279) Total protein (plasma)	[g/L]				84							84	80	95%
Samples and groups														
Sample A		69,0	1,4	2,0	84	CVP	69	0,37	8%	63,4	74,6	84	82	98%
All results		69,0	1,4	2,0	84	0						84		
Sample B		51,5	1,4	2,7	84	CVP	51,5	0,37	8%	47,3	55,7	84	81	96%
All results		51,5	1,4	2,7	84	1						84		
(314) FLC kappa (plasma)	[mg/L]				58							58	56	97%
Samples and groups														
Sample A		1,10	0,44	40	58	CVP	1,1	0,14	±3	0	4,1	58	56	97%
All results		1,10	0,44	40	58	0						58		
Sample B		4,70	1,2	26	58	CVP	4,7	0,40	±3	1,7	7,7	58	58	100%
All results		4,70	1,2	26	58	0						58		
(315) FLC lambda (plasma)	[mg/L]				58							58	49	84%
Samples and groups														
Sample A		149	20	14	58	CVP	149	6,5	50%	74,5	224	58	49	84%
All results		149	20	14	58	0						58		
Sample B		135	24	18	58	CVP	135	7,7	50%	67,5	203	58	49	84%
All results		135	24	18	58	0						58		

st_kn_p

End of report

Printed: 14.03.2022