

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: HS2/22 - Haemocoagulation Special

Deadline: 18.11.2022

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Set 3

(488) APC - resistance

A3	(38) IL	>>>	12	Negative	20 %
	(51) Technoclone	>>>	2	Negative	3,3 %
	(94) Stago	>>>	3	Negative	5,0 %
	(102) Chromogenix	>>>	1	Negative	1,7 %
	(149) Siemens (Dade, BN, Dimension)	>>>	11	Negative	18 %
	(179) Siemens	>>>	3	Negative	5,0 %
	(192) Pentapharm	>>>	27	Negative	45 %
	(193) HYPHEN BioMed	>>>	1	Negative	1,7 %

----- Total per sample: n = 60 ----- Success: 100 % -----

Total success: 100 %

st_kl

End of report

Printed: 23.11.2022