

## Summary statistics - qualitative results

(Groups: manufacturer of instrument)

Deadline: 11.04.2022

## EQA round: IFT1/22 - Immunophenotypisation

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

## Test

Sample	Group	N	Result	N <sub>rel</sub>
<b>(259) HLA-B27</b>				
A	(12) Beckman Coulter	2	Negative	3,2 %
		>	Inconclusive result (borderline)	3,2 %
		>>>	Positive	68 %
	(88) Becton Dickinson	>>>	10 Positive	16 %
	(177) Mindray		1 Negative	1,6 %
		>>>	5 Positive	7,9 %
----- Total per sample: n = 63		----- Success: 95 %	-----	
B	(12) Beckman Coulter	>>>	47 Negative	75 %
	(88) Becton Dickinson	>>>	10 Negative	16 %
	(177) Mindray	>>>	6 Negative	9,5 %
----- Total per sample: n = 63		----- Success: 100 %	-----	
Total success: 95 %				

st\_kl

End of report

Printed: 14.04.2022