

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: IFT2/22 - Immunophenotypisation

Deadline: 17.10.2022

Key: >>> ... marks correct (expected) result

N ... number of the results

> ... marks conditionally correct (acceptable) result

Nrel ... relative number of the results

± ... marks the result not evaluated

Test

Sample	Group	N	Result	N _{rel}
(259) HLA-B27				
A	(0) not specified	>>>	1 Negative	1,5 %
	(12) Beckman Coulter	>>>	46 Negative	68 %
	(36) Immunotech	>>>	2 Negative	2,9 %
	(88) Becton Dickinson	>>>	15 Negative	22 %
	(177) Mindray	>>>	3 Negative	4,4 %
	(190) Exbio Praha	>>>	1 Positive	1,5 %
Total per sample: n = 68		Success:	99 %	
B	(0) not specified	>>>	1 Positive	1,5 %
	(12) Beckman Coulter	>>>	1 Negative	1,5 %
		>	1 Inconclusive result (borderline)	1,5 %
	(36) Immunotech	>>>	44 Positive	65 %
	(88) Becton Dickinson	>>>	2 Positive	2,9 %
		>	1 Inconclusive result (borderline)	1,5 %
	(177) Mindray	>>>	14 Positive	21 %
	(190) Exbio Praha	>>>	3 Positive	4,4 %
Total per sample: n = 68		Success:	99 %	

Total success: 97 %

st_kl

End of report

Printed: 21.10.2022