

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability								
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}	
Set 1															
(536) Prothrombin test (CoaguChek)					145							145	144	99%	
Samples and groups															
Sample A1		1,24	0,056	4,5	145		CVP	1,24	0,011	20%	0,992	1,49	145	144	99%
All results		1,24	0,056	4,5	145	0							145		
Sample B1		2,73	0,11	3,9	145		CVP	2,73	0,021	20%	2,18	3,28	145	145	100%
All results		2,73	0,11	3,9	145	0							145		
Set 3															
(543) Prothrombin test (microINR)					31								31	18	58%
Samples and groups															
Sample A3		2,73	0,43	16	31		CVP	2,73	0,19	20%	2,18	3,28	31	25	81%
All results		2,73	0,43	16	31	0							31		
Sample B3		1,87	0,16	8,7	31		CVP	1,87	0,074	20%	1,49	2,25	31	24	77%
All results		1,87	0,16	8,7	31	2							31		
Set 5															
(552) Prothrombin test (Xprecia)					15								15	15	100%
Samples and groups															
Sample A5		1,68	0,075	4,4	15		CVP	1,68	0,047	20%	1,34	2,02	15	15	100%
All results		1,68	0,075	4,4	15	0							15		
Sample B5		2,57	0,12	4,6	15		CVP	2,57	0,076	20%	2,05	3,09	15	15	100%
All results		2,57	0,12	4,6	15	0							15		

st_kn_p

End of report

Printed: 09.03.2022