

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set 1														
(536) Prothrombin test (CoaguChek)					94						94	94	100%	
Samples and groups	[-]													
Sample A1		2,46	0,12	4,9	94	CVP	2,46	0,031	20%	1,96	2,96	94	94	100%
All results		2,46	0,12	4,9	94	0						94		
Sample B1		2,00	0,092	4,6	94	CVP	2,00	0,023	20%	1,6	2,4	94	94	100%
All results		2,00	0,092	4,6	94	0						94		
st_kn_p														

End of report

Printed: 21.06.2022