

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of the groups n = 5

Deadline: 29.07.2022

## EQA round: KD3/22 - Glycated Haemoglobin

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of the participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Traceability													
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>						
<b>(220) Haemoglobin A1c</b>	[mmol/mol]				17						0				17	16	94%			
Samples and groups																				
<b>Sample A</b>		46,2	2,9	6,2	17						0	CRV	45,5	1,0	13%	39,5	51,5	17	17	100%
(1) HPLC, LC		45,5	2,4	5,2	14	0												14		
Other					3	0												3		
-----												3x4								
<b>Sample B</b>		65,6	4,5	6,8	17						0	CRV	65,9	1,0	13%	57,3	74,5	17	16	94%
(1) HPLC, LC		64,2	2,6	4,1	14	0												14		
Other					3	0												3		
												3x4								

st\_kn\_np

End of report

Printed: 05.08.2022