

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(1001) DAT (direct Coombs test)				
1	All results	>	5 Positive (2+)	5,8 %
		>>>	66 Positive (3+)	77 %
		>	15 Positive (4+)	17 %
----- Total per sample: n = 86 -----		Success: 100 %		
2	All results	>	3 Positive (1+)	3,5 %
		>>>	45 Positive (2+)	52 %
		>	36 Positive (3+)	42 %
		>	2 Positive (4+)	2,3 %
----- Total per sample: n = 86 -----		Success: 100 %		
Total success: 100 %				

(1002) Sensitization specificity anti-IgG

1	All results	>	1 Positive (2+)	2,4 %
		>>>	29 Positive (3+)	69 %
		>	12 Positive (4+)	29 %
----- Total per sample: n = 42 -----		Success: 100 %		
2	All results	>	1 Positive (1+)	2,4 %
		>	22 Positive (2+)	52 %
		>>>	18 Positive (3+)	43 %
		>	1 Positive (4+)	2,4 %
----- Total per sample: n = 42 -----		Success: 100 %		
Total success: 100 %				

(1003) Sensitization specificity anti-IgA

1	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
2	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
Total success: 100 %				

(1004) Sensitization specificity anti-IgM

1	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
2	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
Total success: 100 %				

(1005) Sensitization specificity anti-C3c

1	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
2	All results	>>>	31 Negative	100 %
----- Total per sample: n = 31 -----		Success: 100 %		
Total success: 100 %				

(1006) Sensitization specificity anti-C3d

1	All results	>>>	37 Negative	97 %
			1 Positive (4+)	2,6 %
----- Total per sample: n = 38 -----		Success: 97 %		
2	All results	>>>	36 Negative	95 %
			1 Positive (2+)	2,6 %
			1 Positive (3+)	2,6 %
----- Total per sample: n = 38 -----		Success: 95 %		
Total success: 95 %				

(1008) Col. agglut: Quantif. IgG-DAT

1	All results	>	2 1:30	9,1 %
		>>>	20 1:100	91 %
----- Total per sample: n = 22 -----		Success: 100 %		
2	All results	>>>	21 1:30	95 %
		>	1 1:100	4,5 %
----- Total per sample: n = 22 -----		Success: 100 %		
Total success: 100 %				

(1009) Col. agglut: Subclass IgG1 1:1

1	All results		1 Negative	4,3 %
		>	4 Positive (1+)	17 %
		>>>	15 Positive (2+)	65 %
		>	3 Positive (3+)	13 %
----- Total per sample: n = 23 -----		Success: 96 %		
2	All results	>>>	22 Negative	96 %

EQA round: PAT2/22 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
	2	All results	1	Positive (2+)	4,3 %
	Total per sample: n = 23		Success: 96 %		
	Total success: 96 %				
(1010) Col. agglut: Subclass IgG1 1:100					
	1	All results	>>> 22	Negative	96 %
			1	Positive (1+)	4,3 %
	Total per sample: n = 23		Success: 96 %		
	2	All results	>>> 23	Negative	100 %
	Total per sample: n = 23		Success: 100 %		
	Total success: 96 %				
(1011) Col. agglut: Subclass IgG3 1:1					
	1	All results	>>> 23	Negative	100 %
	Total per sample: n = 23		Success: 100 %		
	2	All results	>>> 23	Negative	100 %
	Total per sample: n = 23		Success: 100 %		
	Total success: 100 %				
(1012) Col. agglut: Subclass IgG3 1:100					
	1	All results	>>> 23	Negative	100 %
	Total per sample: n = 23		Success: 100 %		
	2	All results	>>> 23	Negative	100 %
	Total per sample: n = 23		Success: 100 %		
	Total success: 100 %				
(1019) AB0 blood group					
	1	All results	>>> 56	Blood group: AB	100 %
	Total per sample: n = 56		Success: 100 %		
	2	All results	>>> 55	Blood group: 0	98 %
			1	Not specified	1,8 %
	Total per sample: n = 56		Success: 98 %		
	Total success: 98 %				
(1020) Rh D antigen					
	1	All results	>>> 57	Positive (RhD+)	100 %
	Total per sample: n = 57		Success: 100 %		
	2	All results	>>> 57	Positive (RhD+)	100 %
	Total per sample: n = 57		Success: 100 %		
	Total success: 100 %				
(1021) Antigen C					
	1	All results	>>> 47	Positive	100 %
	Total per sample: n = 47		Success: 100 %		
	2	All results	>>> 47	Negative	100 %
	Total per sample: n = 47		Success: 100 %		
	Total success: 100 %				
(1022) Antigen c					
	1	All results	>>> 47	Positive	100 %
	Total per sample: n = 47		Success: 100 %		
	2	All results	>>> 47	Positive	100 %
	Total per sample: n = 47		Success: 100 %		
	Total success: 100 %				
(1023) Antigen Cw					
	1	All results	>>> 40	Negative	100 %
	Total per sample: n = 40		Success: 100 %		
	2	All results	>>> 40	Negative	100 %
	Total per sample: n = 40		Success: 100 %		
	Total success: 100 %				
(1024) Antigen E					
	1	All results	>>> 48	Negative	100 %
	Total per sample: n = 48		Success: 100 %		
	2	All results	>>> 48	Positive	100 %
	Total per sample: n = 48		Success: 100 %		
	Total success: 100 %				
(1025) Antigen e					
	1	All results	>>> 47	Positive	100 %
	Total per sample: n = 47		Success: 100 %		
	2	All results	>>> 47	Positive	100 %
	Total per sample: n = 47		Success: 100 %		
	Total success: 100 %				

EQA round: PAT2/22 - Direct Antiglobulin Test

Test	Sample	Group	N	Result	N _{rel}
(1026) Antigen K					
1		All results	>>> 49	Negative	100 %
		————— Total per sample: n = 49	————— Success: 100 %		
2		All results	>>> 49	Negative	100 %
		————— Total per sample: n = 49	————— Success: 100 %		
		Total success: 100 %			
(1027) Antigen k					
1		All results	>>> 26	Positive	100 %
		————— Total per sample: n = 26	————— Success: 100 %		
2		All results	>>> 26	Positive	100 %
		————— Total per sample: n = 26	————— Success: 100 %		
		Total success: 100 %			
(1028) Antigen M					
1		All results	>>> 25	Negative	100 %
		————— Total per sample: n = 25	————— Success: 100 %		
2		All results	>>> 25	Positive	100 %
		————— Total per sample: n = 25	————— Success: 100 %		
		Total success: 100 %			
(1029) Antigen N					
1		All results	>>> 23	Positive	100 %
		————— Total per sample: n = 23	————— Success: 100 %		
2		All results	>>> 23	Positive	100 %
		————— Total per sample: n = 23	————— Success: 100 %		
		Total success: 100 %			
(1030) Antigen S					
1		All results	>>> 25	Positive	100 %
		————— Total per sample: n = 25	————— Success: 100 %		
2		All results	>>> 25	Positive	100 %
		————— Total per sample: n = 25	————— Success: 100 %		
		Total success: 100 %			
(1031) Antigen s					
1		All results	>>> 19	Positive	100 %
		————— Total per sample: n = 19	————— Success: 100 %		
2		All results	>>> 18	Negative	95 %
			1	Positive	5,3 %
		————— Total per sample: n = 19	————— Success: 95 %		
		Total success: 95 %			
(1032) Antigen P1					
1		All results	1	Negative	4,5 %
			>>> 21	Positive	95 %
		————— Total per sample: n = 22	————— Success: 95 %		
2		All results	>>> 19	Negative	86 %
			3	Positive	14 %
		————— Total per sample: n = 22	————— Success: 86 %		
		Total success: 82 %			
(1033) Antigen Jka					
1		All results	>>> 24	Negative	92 %
			2	Positive	7,7 %
		————— Total per sample: n = 26	————— Success: 92 %		
2		All results	>>> 26	Positive	100 %
		————— Total per sample: n = 26	————— Success: 100 %		
		Total success: 92 %			
(1034) Antigen Jkb					
1		All results	>>> 26	Positive	100 %
		————— Total per sample: n = 26	————— Success: 100 %		
2		All results	>>> 26	Positive	100 %
		————— Total per sample: n = 26	————— Success: 100 %		
		Total success: 100 %			
(1035) Antigen Lea					
1		All results	>>> 19	Negative	95 %
			1	Positive	5,0 %
		————— Total per sample: n = 20	————— Success: 95 %		
2		All results	>>> 18	Negative	90 %
			2	Positive	10 %
		————— Total per sample: n = 20	————— Success: 90 %		
		Total success: 90 %			

Test

Sample	Group	N	Result	N _{rel}
(1036) Antigen Leb				
1	All results	>>> 20	Positive	100 %
----- Total per sample: n = 20		----- Success: 100 %	-----	
2	All results	>>> 20	Positive	100 %
----- Total per sample: n = 20		----- Success: 100 %	-----	
Total success: 100 %				

st_kl

End of report

Printed: 26.11.2022